

Combined uncertainty in the theoretical values (%) for the EMEP laboratory intercomparison Round 41

Sample type G	H+	SO4-S	NO3-N	Na	NH4-N	Cl	Mg	Ca	K	pH	Cond.
G1	0.31	0.55	0.23	0.21	0.14	0.17	0.23	0.40	0.14	0.31	0.30
G2	0.29	0.62	0.86	0.25	0.83	0.20	0.23	0.50	0.13	0.29	0.50
G3	0.57	0.75	1.00	0.25	0.98	0.19	0.23	0.43	0.13	0.57	0.58
G4	0.59	0.79	0.26	0.22	0.17	0.16	0.23	0.46	0.14	0.59	0.32

Sample type H	As	Cd	Cr	Cu	NI	Pb	Zn
H1	0.59	0.76	0.57	0.52	0.68	0.52	0.61
H2	1.00	0.83	0.57	0.59	1.00	0.54	0.92
H3	0.72	0.65	1.09	0.84	0.89	0.58	0.65
H4	1.08	0.72	1.08	0.73	0.80	0.77	0.62

Sample type B	SO2	HNO3
B1	2.0	2.0
B2	1.7	1.7
B4	1.3	1.3
B6	1.4	1.3

Sample type C	NO2
C1	0.29
C2	0.25
C3	0.19
C4	0.38

Sample type J	NH3
J1	2.5
J2	1.8
J4	1.4
J5	1.3