

Lab	Lab name	Component	QA measure ID	QA date	QA document url	QA bias	QA variability	QA outcome
1	AT	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.90%	2.97%	Pass
1	AT	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.42% S	1.57%	Pass
1	AT	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-12.95% S	1.31%	Pass
1	AT	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
1	AT	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.11% S	0.57%	Pass
1	AT	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.54% S	0.54%	Pass
1	AT	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
1	AT	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.72% S	3.04%	Pass
1	AT	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.64% S	1.88%	Pass
1	AT	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.77% S	0.46%	Pass
2	BE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.34%	0.67%	Pass
2	BE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.29%	2.47%	Pass
2	BE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.88% S	0.21%	Pass
2	BE	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.44% S	0.14%	Pass
2	BE	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.26%	0.56%	Pass
2	BE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.33% S	0.21%	Pass
2	BE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	19.57% S	0.64%	Pass
3	CZ	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.90%	5.94%	Pass
3	CZ	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.49%	2.54%	Pass
3	CZ	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	8.77%	8.78%	Pass
3	CZ	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.26%	2.35%	Pass
3	CZ	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.17%	2.19%	Pass
3	CZ	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.12% S	4.80%	Pass
3	CZ	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.71% S	2.93%	Pass
3	CZ	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.04%	0.53%	Pass
3	CZ	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	77.72% S	99.14%	No pass
3	CZ	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.96% S	0.39%	Pass
3	CZ	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00%	0.29%	Pass
3	CZ	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.09%	3.38%	Pass
3	CZ	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.44% S	1.30%	Pass
3	CZ	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.76% S	0.54%	Pass
3	CZ	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-11.99% S	4.24%	
3	CZ	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.66% S	0.80%	Pass
3	CZ	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-39.58%	47.67%	No pass
3	CZ	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.69%	1.88%	Pass
3	CZ	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.35% S	0.42%	Pass
3	CZ	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	16.77% S	12.51%	Pass
4	DK	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.65%	1.65%	Pass
4	DK	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.22%	3.35%	Pass
4	DK	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.43%	19.79%	Pass

4	DK	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.82%	18.26%	No pass
4	DK	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.48% S	0.29%	Pass
4	DK	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.85% S	2.11%	Pass
4	DK	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
4	DK	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.83% S	0.00%	Pass
4	DK	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.58% S	0.67%	Pass
4	DK	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.78% S	1.50%	Pass
4	DK	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.92% S	2.59%	Pass
4	DK	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.37%	6.01%	Pass
4	DK	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.50% S	0.00%	Pass
4	DK	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.29%	0.72%	Pass
4	DK	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.67% S	5.24%	
4	DK	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.52%	1.60%	Pass
4	DK	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.40% S	2.56%	Pass
4	DK	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.98% S	0.99%	Pass
4	DK	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.94%	0.69%	Pass
4	DK	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
5	FI	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.23%	2.15%	Pass
5	FI	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.84% S	2.51%	Pass
5	FI	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	8.64% S	5.57%	Pass
5	FI	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.95% S	0.26%	Pass
5	FI	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.07%	2.48%	Pass
5	FI	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.86% S	2.21%	Pass
5	FI	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
5	FI	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.89% S	3.14%	Pass
5	FI	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.11%	1.12%	Pass
5	FI	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	8.59% S	6.36%	Pass
5	FI	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.21% S	1.15%	Pass
5	FI	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.08% S	11.17%	No pass
5	FI	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.22% S	3.04%	Pass
5	FI	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.59% S	1.71%	Pass
5	FI	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.50% S	1.31%	
5	FI	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	8.33% S	4.32%	Pass
5	FI	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.28%	1.72%	Pass
5	FI	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.69% S	0.77%	Pass
5	FI	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.75% S	0.85%	Pass
5	FI	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	8.16% S	5.29%	Pass
6	COM	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	29.61% S	4.95%	Pass
6	COM	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.06%	7.29%	Pass
6	COM	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.00%	1.24%	Pass
6	COM	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-15.59% S	2.09%	Pass

6	COM	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.63% S	0.88%	Pass
6	COM	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.26% S	1.79%	Pass
6	COM	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.19%	0.77%	Pass
6	COM	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-15.56% S	0.71%	Pass
6	COM	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-46.72% S	25.75%	No pass
6	COM	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.70% S	1.35%	Pass
6	COM	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-15.32% S	1.15%	Pass
6	COM	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	25.58% S	13.59%	No pass
6	COM	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.42% S	2.41%	Pass
6	COM	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.50% S	0.81%	Pass
6	COM	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.17%	12.84%	
6	COM	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.47% S	1.12%	Pass
6	COM	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.48%	24.46%	No pass
6	COM	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.25% S	1.44%	Pass
6	COM	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.83% S	0.46%	Pass
6	COM	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	17.83% S	3.19%	Pass
7	COM	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.40%	3.35%	Pass
7	COM	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00% S	0.00%	Pass
7	COM	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	25.17% S	18.44%	Pass
7	COM	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	23.06% S	16.04%	Pass
7	COM	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.77% S	3.57%	Pass
7	COM	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00%	3.66%	Pass
7	COM	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	29.43% S	19.05%	Pass
8	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.23%	2.97%	Pass
8	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.33% S	1.40%	Pass
8	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.03% S	6.56%	Pass
8	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-15.47% S	2.09%	Pass
8	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.40% S	0.88%	Pass
8	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.25% S	4.32%	Pass
8	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.50%	0.64%	Pass
8	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.92% S	4.92%	Pass
8	DE	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.63%	2.66%	Pass
8	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.67% S	8.35%	Pass
8	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.05% S	1.15%	Pass
8	DE	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.21% S	0.99%	Pass
8	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.55% S	5.34%	Pass
8	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.09%	5.77%	No pass
8	DE	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.09% S	5.05%	Pass
8	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-10.87% S	3.18%	
8	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.92% S	1.44%	Pass
8	DE	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.80%	1.73%	Pass

8	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.93%	1.10%	Pass
8	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.78% S	1.15%	Pass
8	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.38% S	3.87%	Pass
10	HU	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.94% S	5.12%	Pass
10	HU	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
10	HU	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-28.87% S	12.00%	Pass
10	HU	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.02%	9.65%	Pass
10	HU	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.98%	9.34%	Pass
10	HU	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
10	HU	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.78% S	1.36%	Pass
10	HU	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
10	HU	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	23.01% S	8.46%	Pass
10	HU	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.40%	1.44%	Pass
10	HU	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-30.40% S	15.90%	No pass
10	HU	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
10	HU	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.06%	3.88%	Pass
10	HU	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.47% S	0.39%	Pass
10	HU	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-13.89% S	3.96%	
10	HU	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.14% S	1.44%	Pass
10	HU	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-11.15% S	1.66%	Pass
10	HU	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.94%	5.08%	No pass
10	HU	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
12	IE	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.43%	3.89%	Pass
13	IT	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.78%	1.65%	Pass
13	IT	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.59% S	2.09%	Pass
13	IT	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.13%	1.75%	Pass
13	IT	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.33%	2.83%	Pass
13	IT	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.55% S	1.72%	Pass
13	IT	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.81%	5.66%	Pass
13	IT	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.52%	1.71%	Pass
13	IT	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.67%	3.52%	
13	IT	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.06% S	0.64%	Pass
13	IT	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.21%	0.66%	Pass
13	IT	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.29%	3.54%	Pass
15	NO	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	20.13% S	9.41%	No pass
15	NO	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.53%	3.35%	Pass
15	NO	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00% S	1.24%	Pass
15	NO	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.13% S	5.48%	Pass
15	NO	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.05% S	1.02%	Pass
15	NO	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.67%	1.58%	Pass
15	NO	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.98% S	1.09%	Pass

15	NO	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.42%	2.28%	Pass
15	NO	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.44%	4.95%	Pass
15	NO	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.26%	2.25%	Pass
15	NO	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	15.79%	15.81%	No pass
15	NO	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.83%	2.86%	Pass
15	NO	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.67%	4.50%	Pass
15	NO	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.52%	1.80%	Pass
15	NO	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-12.88% S	5.83%	Pass
15	NO	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-15.85% S	6.58%	
15	NO	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.67%	4.16%	Pass
15	NO	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	8.33%	4.16%	Pass
15	NO	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.56%	7.61%	Pass
15	NO	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.46% S	0.44%	Pass
15	NO	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.03%	0.65%	Pass
15	NO	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.65% S	3.27%	Pass
16	PL	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.87%	494.69%	No pass
16	PL	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
16	PL	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00% S	0.62%	Pass
16	PL	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.03%	2.09%	Pass
16	PL	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.00%	4.38%	Pass
16	PL	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00% S	0.00%	Pass
16	PL	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.04% S	0.64%	Pass
16	PL	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.08%	3.57%	Pass
16	PL	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.38%	1.61%	Pass
16	PL	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00%	3.94%	Pass
16	PL	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.66% S	0.57%	Pass
16	PL	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.59% S	2.62%	Pass
16	PL	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.67% S	1.05%	Pass
16	PL	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.58% S	1.71%	Pass
16	PL	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.34%	1.56%	Pass
16	PL	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.22%	4.80%	
16	PL	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.73%	0.80%	Pass
16	PL	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.04%	3.93%	Pass
16	PL	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.19%	2.10%	Pass
16	PL	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.65%	1.04%	Pass
16	PL	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.95% S	3.17%	Pass
18	RO	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
18	RO	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
18	RO	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.14%	0.44%	Pass
18	RO	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
18	RO	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			

18	RO	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.69% S	0.45%	Pass
18	RO	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
18	RO	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
18	RO	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
18	RO	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.41% S	0.31%	Pass
19	ES	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	16.85% S	4.29%	Pass
19	ES	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.50%	2.09%	Pass
19	ES	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.82% S	1.31%	Pass
19	ES	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.10% S	0.93%	Pass
19	ES	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.03% S	0.29%	Pass
19	ES	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.95%	1.19%	Pass
19	ES	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.11%	7.94%	No pass
19	ES	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.45%	3.50%	Pass
19	ES	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-16.82% S	4.49%	
19	ES	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	11.71% S	6.72%	Pass
19	ES	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-15.04% S	4.16%	Pass
19	ES	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.38% S	6.96%	Pass
19	ES	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.69%	6.58%	No pass
20	SE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.94% S	1.49%	Pass
20	SE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.57%	1.25%	Pass
20	SE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.30%	4.95%	Pass
20	SE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.43% S	1.57%	Pass
20	SE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.05% S	2.04%	Pass
20	SE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.54%	1.68%	Pass
20	SE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.01%	0.80%	Pass
20	SE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.53%	1.50%	Pass
20	SE	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.94%	2.69%	Pass
20	SE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.22%	3.04%	Pass
20	SE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.74% S	0.86%	Pass
20	SE	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.66%	0.90%	Pass
20	SE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.83%	2.34%	Pass
20	SE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.04%	1.26%	Pass
20	SE	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.69%	1.56%	Pass
20	SE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.50% S	1.40%	
20	SE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.68% S	0.80%	Pass
20	SE	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.31%	7.63%	Pass
20	SE	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.48%	0.71%	Pass
20	SE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.52% S	0.66%	Pass
20	SE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.66%	0.58%	Pass
20	SE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.45%	0.87%	Pass
21	CH	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	12.02% S	5.61%	Pass

21	CH	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.52% S	1.83%	Pass
21	CH	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.01%	2.92%	Pass
21	CH	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.46%	0.30%	Pass
21	CH	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.87% S	2.30%	Pass
21	CH	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.19%	0.27%	Pass
21	CH	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-8.80% S	4.17%	
21	CH	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.38% S	5.76%	Pass
21	CH	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.28% S	0.55%	Pass
21	CH	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.98% S	0.85%	Pass
22	RU	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	11.79% S	3.47%	Pass
22	RU	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	45.36% S	6.52%	Pass
22	RU	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-10.01%	10.94%	No pass
22	RU	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.53% S	3.54%	Pass
22	RU	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	14.72% S	11.19%	No pass
22	RU	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.50% S	2.59%	Pass
22	RU	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.84%	1.08%	Pass
22	RU	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.89% S	5.11%	
22	RU	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.60%	4.48%	Pass
22	RU	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.09% S	15.42%	No pass
22	RU	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-8.53%	8.73%	Pass
22	RU	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.11% S	0.38%	Pass
23	COM	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.75% S	1.98%	Pass
23	COM	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-11.02% S	3.52%	Pass
23	COM	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.32% S	1.02%	Pass
23	COM	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.28% S	1.45%	Pass
23	COM	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.86%	0.86%	Pass
23	COM	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.56% S	0.68%	Pass
23	COM	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.81% S	0.93%	Pass
23	COM	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.28% S	11.64%	
23	COM	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.71% S	2.48%	Pass
23	COM	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.12%	1.33%	Pass
23	COM	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.67% S	1.00%	Pass
24	RS	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.94% S	1.65%	Pass
24	RS	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.22%	6.28%	Pass
24	RS	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-17.67%	3.71%	Pass
24	RS	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.39% S	1.57%	Pass
24	RS	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.23% S	2.19%	Pass
24	RS	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.45% S	12.33%	Pass
24	RS	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.90% S	0.32%	Pass
24	RS	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	23.33% S	12.69%	Pass
24	RS	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-45.26% S	12.01%	Pass

24	RS	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.05%	2.87%	Pass
24	RS	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-21.33% S	9.42%	Pass
24	RS	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.31% S	1.35%	Pass
24	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.11% S	1.94%	Pass
24	RS	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
24	RS	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.64% S	2.08%	Pass
24	RS	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.68%	0.99%	Pass
24	RS	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.98%	1.31%	Pass
24	RS	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
26	CA	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.14% S	0.50%	Pass
26	CA	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.19% S	1.30%	Pass
26	CA	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.29%	0.73%	Pass
26	CA	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
26	CA	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.97% S	2.01%	Pass
26	CA	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.74% S	0.27%	Pass
26	CA	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.74% S	0.97%	
26	CA	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.24% S	1.60%	Pass
26	CA	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.07% S	0.66%	Pass
26	CA	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.04%	0.12%	Pass
27	EDU	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.33%	0.33%	Pass
27	EDU	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.35%	1.33%	Pass
27	EDU	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.09%	0.79%	Pass
27	EDU	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.59% S	0.20%	Pass
27	EDU	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.34% S	0.57%	Pass
27	EDU	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.35%	0.74%	Pass
27	EDU	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.84% S	4.74%	
27	EDU	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.07% S	2.71%	Pass
27	EDU	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.39%	1.97%	Pass
27	EDU	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.24% S	0.15%	Pass
30	EU	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.48%	2.31%	Pass
30	EU	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.68% S	2.09%	Pass
30	EU	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.70% S	1.61%	Pass
30	EU	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.26%	1.39%	Pass
30	EU	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.66% S	1.15%	Pass
30	EU	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.75% S	1.35%	Pass
30	EU	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	17.52% S	3.68%	
30	EU	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.07%	7.04%	Pass
30	EU	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.34% S	0.44%	Pass
30	EU	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.11% S	2.81%	Pass
31	SK	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.69%	5.28%	Pass
31	SK	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.28% S	2.95%	Pass

31	SK	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-11.67% S	7.67%	Pass
31	SK	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.69%	1.57%	Pass
31	SK	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.06%	7.59%	Pass
31	SK	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.43% S	2.31%	Pass
31	SK	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.60%	0.83%	Pass
31	SK	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.61% S	6.38%	Pass
31	SK	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.90%	1.72%	Pass
31	SK	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-12.36% S	6.33%	Pass
31	SK	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.73% S	0.29%	Pass
31	SK	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.02%	4.62%	Pass
31	SK	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	4.52% S	1.84%	Pass
31	SK	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.37% S	0.27%	Pass
31	SK	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.73% S	0.78%	Pass
31	SK	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-10.87% S	4.22%	
31	SK	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.51% S	1.44%	Pass
31	SK	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.29%	3.11%	Pass
31	SK	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.26%	1.99%	Pass
31	SK	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.91% S	0.73%	Pass
31	SK	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.63%	2.12%	Pass
32	LT	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.72% S	1.49%	Pass
32	LT	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.49% S	8.37%	Pass
32	LT	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00% S	0.00%	Pass
32	LT	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-50.48% S	22.17%	No pass
32	LT	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.15%	1.75%	Pass
32	LT	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-17.14% S	14.75%	Pass
32	LT	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.44% S	1.38%	Pass
32	LT	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.56%	13.55%	Pass
32	LT	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.29%	2.42%	Pass
32	LT	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-20.00% S	14.27%	Pass
32	LT	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
32	LT	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.97%	2.51%	Pass
32	LT	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	4.17%	16.75%	Pass
32	LT	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.07% S	2.80%	Pass
32	LT	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.23% S	1.94%	Pass
32	LT	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-13.89% S	5.58%	
32	LT	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.18% S	0.80%	Pass
32	LT	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.49%	34.16%	No pass
32	LT	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.84%	2.10%	Pass
32	LT	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.75%	1.12%	Pass
32	LT	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	9.45%	21.16%	Pass
33	LV	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.80%	0.83%	Pass

33	LV	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.95%	40.20%	No pass
33	LV	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00%	7.42%	Pass
33	LV	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.65%	2.61%	Pass
33	LV	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	20.28% S	6.86%	Pass
33	LV	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	13.83% S	1.05%	Pass
33	LV	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.19% S	0.58%	Pass
33	LV	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00%	5.70%	Pass
33	LV	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.24% S	2.68%	Pass
33	LV	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-11.14%	5.63%	Pass
33	LV	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.47% S	1.15%	Pass
33	LV	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.13% S	1.71%	Pass
33	LV	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.78%	36.64%	No pass
33	LV	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	9.81% S	1.35%	Pass
33	LV	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.89%	1.63%	Pass
33	LV	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-12.90% S	3.77%	
33	LV	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.36%	4.32%	Pass
33	LV	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.60% S	3.11%	Pass
33	LV	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.47% S	4.42%	Pass
33	LV	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.28% S	3.58%	Pass
33	LV	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00%	2.12%	Pass
34	TR	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	56.13% S	28.23%	No pass
34	TR	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-22.35% S	13.44%	Pass
34	TR	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	42.86%	0.00%	Pass
34	TR	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	29.63%	22.96%	No pass
34	TR	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	851.39% S	569.14%	No pass
34	TR	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-33.45% S	20.72%	Pass
34	TR	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-11.27% S	3.10%	Pass
34	TR	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.12%	13.48%	Pass
34	TR	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-99.39%	68.71%	No pass
34	TR	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	132.46% S	9.39%	Pass
34	TR	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-41.49% S	7.76%	Pass
34	TR	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-95.70%	59.39%	No pass
34	TR	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-21.09% S	12.59%	Pass
34	TR	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.71% S	2.07%	Pass
34	TR	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	66.03% S	18.27%	No pass
34	TR	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-22.69% S	10.93%	
34	TR	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-21.21%	155.45%	No pass
34	TR	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-99.30%	78.43%	No pass
34	TR	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.62%	9.17%	Pass
34	TR	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.67% S	1.23%	Pass
34	TR	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.96%	10.32%	Pass

34	Ty	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.68%	1.53%	Pass
35	HR	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.61% S	0.83%	Pass
35	HR	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.96% S	1.04%	Pass
35	HR	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.01% S	0.73%	Pass
35	HR	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.52%	0.93%	Pass
35	HR	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.75% S	0.86%	Pass
35	HR	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.46% S	1.08%	Pass
35	HR	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.74% S	1.79%	
35	HR	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	4.90% S	0.80%	Pass
35	HR	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.37% S	0.55%	Pass
35	HR	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.15% S	1.58%	Pass
36	SI	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.06% S	0.83%	Pass
36	SI	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.73%	0.50%	Pass
36	SI	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	11.60%	5.07%	Pass
36	SI	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.90% S	2.61%	Pass
36	SI	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.38%	4.23%	Pass
36	SI	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.07%	3.57%	Pass
36	SI	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.57% S	0.43%	Pass
36	SI	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	11.56%	11.41%	Pass
36	SI	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.19%	2.38%	Pass
36	SI	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.86%	6.64%	Pass
36	SI	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.44% S	1.72%	Pass
36	SI	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.29% S	1.02%	Pass
36	SI	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.43% S	2.45%	Pass
36	SI	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.62% S	1.98%	Pass
36	SI	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.45% S	1.56%	Pass
36	SI	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.16% S	1.37%	
36	SI	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.65% S	3.20%	Pass
36	SI	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	28.65%	5.27%	Pass
36	SI	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.47% S	0.98%	Pass
36	SI	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.45% S	0.88%	Pass
36	SI	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.11% S	1.04%	Pass
36	SI	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	8.74% S	9.56%	Pass
38	EE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	26.80% S	14.86%	No pass
38	EE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.03%	3.27%	Pass
38	EE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.43%	2.47%	Pass
38	EE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-10.15%	36.26%	No pass
38	EE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.26% S	3.50%	Pass
38	EE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.62% S	3.06%	Pass
38	EE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.53%	30.94%	No pass
38	EE	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.89% S	0.71%	Pass

38	EE	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.76%	4.17%	Pass
38	EE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.24%	1.58%	Pass
38	EE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	13.13%	10.06%	No pass
38	EE	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.80%	6.92%	Pass
38	EE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.42% S	3.45%	Pass
38	EE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	8.90% S	4.33%	Pass
38	EE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.39% S	4.49%	
38	EE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-10.21% S	0.80%	Pass
38	EE	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	96.05% S	70.75%	No pass
38	EE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.03% S	1.33%	Pass
38	EE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.95% S	2.39%	Pass
38	EE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.37% S	3.97%	Pass
39	PL	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	24.93% S	27.57%	No pass
39	PL	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00%	0.78%	Pass
39	PL	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.20% S	1.61%	Pass
39	PL	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
39	PL	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.67%	4.30%	Pass
39	PL	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.84% S	0.29%	Pass
39	PL	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.50%	5.12%	Pass
39	PL	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.07% S	0.18%	Pass
39	PL	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.90% S	2.33%	Pass
39	PL	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.74%	23.11%	
39	PL	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.77% S	0.80%	Pass
39	PL	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.44%	2.45%	Pass
39	PL	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.60% S	1.22%	Pass
39	PL	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.02% S	1.19%	Pass
40	MK	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-20.63% S	14.53%	No pass
40	MK	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
40	MK	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	77.51% S	12.99%	No pass
40	MK	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.86%	10.37%	No pass
40	MK	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
40	MK	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.60%	4.15%	Pass
40	MK	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	464.73% S	195.74%	
40	MK	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
40	MK	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
40	MK	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
42	COM	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.96%	8.92%	No pass
42	COM	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.26%	11.22%	Pass
42	COM	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-25.24% S	4.67%	Pass
42	COM	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-23.73% S	11.22%	No pass
42	COM	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-28.17%	20.99%	No pass

42	COM	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-45.84% S	10.37%	No pass
42	COM	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-30.23%	28.77%	No pass
42	COM	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	4.71%	10.24%	
42	COM	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-28.68%	48.99%	No pass
42	COM	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.76%	17.46%	No pass
42	COM	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-12.16% S	5.39%	No pass
45	COM	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	8.89% S	3.63%	Pass
45	COM	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-15.89%	19.30%	No pass
45	COM	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.45% S	1.02%	Pass
45	COM	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.61%	0.96%	Pass
45	COM	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.86%	4.89%	Pass
45	COM	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.18%	0.99%	Pass
45	COM	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.31% S	3.09%	
45	COM	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.06%	10.41%	Pass
45	COM	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.98%	5.86%	Pass
45	COM	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.54% S	1.50%	Pass
46	PL	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
46	PL	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
46	PL	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
46	PL	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.41% S	1.30%	Pass
46	PL	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
46	PL	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
46	PL	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.77% S	4.72%	
46	PL	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
46	PL	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
46	PL	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
48	BE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-20.05% S	2.48%	Pass
48	BE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
48	BE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-13.70%	8.06%	Pass
48	BE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
48	BE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
48	BE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.50%	1.31%	Pass
48	BE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
48	BE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
48	BE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
48	BE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.79% S	0.35%	Pass
107	FI	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.06% S	1.49%	Pass
107	FI	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
107	FI	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.40% S	3.79%	Pass
107	FI	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.00% S	1.03%	Pass
107	FI	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			

107	FI	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.21% S	2.52%	Pass
107	FI	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.67% S	1.93%	
107	FI	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
107	FI	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
107	FI	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.21% S	3.66%	Pass
110	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.16%	4.46%	Pass
110	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.97% S	2.18%	Pass
110	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.57% S	4.95%	Pass
110	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-8.08% S	1.83%	Pass
110	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	9.35% S	7.59%	Pass
110	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.76%	0.42%	Pass
110	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.09% S	3.19%	Pass
110	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.67% S	2.00%	Pass
110	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00%	1.20%	Pass
110	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.84% S	0.86%	Pass
110	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.30%	1.26%	Pass
110	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	16.93% S	3.43%	Pass
110	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-20.38% S	4.02%	
110	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.87% S	2.56%	Pass
110	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.98% S	0.99%	Pass
110	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.36% S	4.08%	Pass
110	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.90%	1.06%	Pass
112	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.96% S	1.32%	Pass
112	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
112	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.43% S	2.47%	Pass
112	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	11.82% S	2.09%	Pass
112	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.79% S	0.73%	Pass
112	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.23% S	2.21%	Pass
112	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.06% S	0.64%	Pass
112	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.92%	0.71%	Pass
112	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.11% S	2.40%	Pass
112	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.92% S	2.59%	Pass
112	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.92%	3.35%	Pass
112	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.36% S	1.62%	Pass
112	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.67% S	1.93%	
112	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.43%	3.52%	Pass
112	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.41% S	1.33%	Pass
112	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.56% S	0.23%	Pass
112	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.41% S	1.54%	Pass
113	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.15%	6.44%	Pass
113	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.98% S	5.44%	Pass

113	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-19.45% S	11.42%	Pass
113	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
113	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	16.25% S	4.23%	Pass
113	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.88%	5.44%	Pass
113	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-8.02% S	5.15%	Pass
113	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-33.68% S	1.07%	Pass
113	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.41% S	2.88%	Pass
113	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
113	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.49%	3.56%	Pass
113	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.30% S	1.80%	Pass
113	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.98%	12.90%	
113	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
113	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
113	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.30% S	1.54%	Pass
113	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-27.12% S	1.76%	Pass
114	IT	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.25%	4.62%	Pass
114	IT	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-28.54% S	28.47%	No pass
114	IT	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	33.33%	17.32%	Pass
114	IT	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.17% S	2.09%	Pass
114	IT	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.52% S	1.90%	Pass
114	IT	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00% S	0.00%	Pass
114	IT	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.58%	2.19%	Pass
114	IT	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-21.11% S	23.53%	Pass
114	IT	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00% S	0.00%	Pass
114	IT	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.03% S	0.29%	Pass
114	IT	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00%	2.09%	Pass
114	IT	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.83% S	1.08%	Pass
114	IT	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.84% S	3.19%	
114	IT	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.96% S	0.32%	Pass
114	IT	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.48%	1.88%	Pass
114	IT	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.17% S	0.81%	Pass
114	IT	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.19%	0.35%	Pass
115	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.71% S	3.80%	Pass
115	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.29% S	0.75%	Pass
115	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.30%	0.62%	Pass
115	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.04% S	1.04%	Pass
115	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.89%	3.21%	Pass
115	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.84% S	5.99%	Pass
115	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.27% S	1.73%	Pass
115	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.25%	3.04%	Pass
115	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.80% S	11.24%	Pass

115	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.17% S	0.57%	Pass
115	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.92% S	5.65%	Pass
115	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.06% S	0.81%	Pass
115	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.67% S	2.17%	
115	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.80% S	3.04%	Pass
115	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.72%	1.33%	Pass
115	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.32%	2.04%	Pass
115	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.01% S	0.39%	Pass
116	CH	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	39.66% S	11.39%	No pass
116	CH	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.70%	8.09%	Pass
116	CH	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-14.31%	51.84%	No pass
116	CH	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-19.36% S	6.00%	Pass
116	CH	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.29% S	0.57%	Pass
116	CH	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.58% S	52.52%	No pass
116	CH	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-42.46% S	12.65%	
116	CH	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	4.33% S	1.12%	Pass
116	CH	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.57% S	0.88%	Pass
116	CH	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.45% S	55.06%	No pass
117	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
117	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
117	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.43% S	13.61%	Pass
117	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.91% S	1.30%	Pass
117	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
117	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.64%	7.27%	Pass
117	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.40% S	1.09%	Pass
117	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.72%	2.64%	Pass
117	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.39%	3.72%	Pass
117	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.41% S	0.57%	Pass
117	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	13.42% S	10.57%	Pass
117	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
117	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.59%	3.52%	
117	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.68% S	1.92%	Pass
117	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.24% S	3.09%	Pass
117	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.86% S	1.58%	Pass
117	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.63%	0.95%	Pass
118	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	4.31% S	2.31%	Pass
118	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-22.78% S	15.91%	Pass
118	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-28.00% S	2.47%	Pass
118	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.13%	3.91%	Pass
118	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.95% S	3.36%	Pass
118	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-16.50% S	4.21%	Pass

118	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.97% S	1.36%	Pass
118	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.50%	4.99%	Pass
118	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	203.77%	36.79%	No pass
118	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.93% S	1.15%	Pass
118	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-11.67% S	2.09%	Pass
118	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.60% S	1.53%	Pass
118	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.86% S	1.93%	
118	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.68% S	1.92%	Pass
118	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.42% S	1.99%	Pass
118	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.94% S	1.62%	Pass
118	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.11%	4.30%	Pass
120	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.17% S	1.32%	Pass
120	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.88%	13.40%	Pass
120	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-17.14% S	0.00%	Pass
120	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.96%	3.91%	Pass
120	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	25.45% S	12.26%	No pass
120	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	11.76%	13.70%	Pass
120	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.62% S	1.94%	Pass
120	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	35.56%	14.26%	Pass
120	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-32.89% S	9.39%	Pass
120	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.17% S	2.30%	Pass
120	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-16.67%	15.70%	Pass
120	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.13% S	1.26%	Pass
120	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.59% S	2.22%	
120	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.08% S	3.52%	Pass
120	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.71% S	0.99%	Pass
120	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.18% S	1.39%	Pass
120	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.49% S	0.71%	Pass
124	BE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.19%	3.30%	Pass
124	BE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
124	BE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
124	BE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.09%	1.57%	Pass
124	BE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.02%	6.57%	Pass
124	BE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	8.62%	0.00%	Pass
124	BE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.73% S	1.67%	Pass
124	BE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.56%	3.57%	Pass
124	BE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.04%	9.01%	Pass
124	BE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.50% S	0.86%	Pass
124	BE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.50% S	3.14%	Pass
124	BE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.60%	1.53%	Pass
124	BE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.79% S	8.24%	

124	BE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.94% S	1.28%	Pass
124	BE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.42%	0.99%	Pass
124	BE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.60% S	0.15%	Pass
124	BE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.29% S	2.33%	Pass
125	DE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.39% S	2.15%	Pass
125	DE	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.74% S	6.63%	Pass
125	DE	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.73%	1.98%	Pass
125	DE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-17.96% S	7.83%	Pass
125	DE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.33% S	2.04%	Pass
125	DE	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.07%	2.84%	Pass
125	DE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-11.89% S	2.54%	Pass
125	DE	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.83% S	3.52%	Pass
125	DE	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.83% S	4.39%	Pass
125	DE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-19.21% S	4.31%	Pass
125	DE	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.58%	5.34%	Pass
125	DE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.62%	0.54%	Pass
125	DE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-13.89% S	4.70%	
125	DE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-11.07% S	6.88%	Pass
125	DE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-20.14% S	8.07%	Pass
125	DE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.51% S	1.54%	Pass
125	DE	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.24% S	10.04%	Pass
126	IT	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.05% S	2.81%	Pass
126	IT	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.54%	24.78%	No pass
126	IT	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.19% S	1.17%	Pass
126	IT	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.64% S	0.91%	Pass
126	IT	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-20.53% S	3.16%	Pass
126	IT	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.85%	8.84%	No pass
126	IT	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.14%	5.77%	
126	IT	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-19.93% S	4.48%	Pass
126	IT	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.88% S	2.32%	Pass
126	IT	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.78% S	0.58%	Pass
129	TN	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
129	TN	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
129	TN	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	9880.77% S	150.66%	No pass
129	TN	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
129	TN	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
129	TN	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
129	TN	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1614.29%	1202.70%	No pass
132	CL	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
132	CL	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
132	CL	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.00%	0.00%	Pass

132	CL	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.86% S	3.57%	Pass
132	CL	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.22% S	0.38%	Pass
132	CL	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-8.00%	31.40%	No pass
132	CL	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-20.42% S	1.34%	Pass
140	IT	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.75%	1.16%	Pass
140	IT	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.11% S	1.83%	Pass
140	IT	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.66%	2.92%	Pass
140	IT	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.15% S	0.62%	Pass
140	IT	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.53% S	0.57%	Pass
140	IT	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.40% S	2.34%	Pass
140	IT	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-12.88% S	5.46%	
140	IT	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.16%	0.96%	Pass
140	IT	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.46%	2.10%	Pass
140	IT	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.70% S	1.73%	Pass
141	JP	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
141	JP	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.13%	0.99%	Pass
141	JP	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
141	JP	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
141	JP	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
141	JP	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
141	JP	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
145	EE	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.43%	1.98%	Pass
145	EE	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.07%	16.43%	No pass
145	EE	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.61% S	1.02%	Pass
145	EE	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.00% S	0.65%	Pass
145	EE	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.31% S	1.44%	Pass
145	EE	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.79% S	2.52%	Pass
145	EE	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-12.90% S	3.73%	
145	EE	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.17% S	1.92%	Pass
145	EE	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.73% S	0.99%	Pass
145	EE	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.58% S	0.69%	Pass
146	LU	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.48% S	3.47%	Pass
146	LU	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.12% S	0.60%	Pass
146	LU	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.23% S	0.99%	Pass
146	LU	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	87.90% S	24.26%	No pass
146	LU	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-19.49% S	1.90%	Pass
146	LU	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.93% S	2.31%	Pass
146	LU	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.81% S	1.97%	Pass
146	LU	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.80% S	1.78%	Pass
146	LU	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.43% S	0.83%	Pass
146	LU	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.73%	0.29%	Pass

146	LU	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-11.32% S	2.55%	Pass
146	LU	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	13.32% S	3.88%	Pass
146	LU	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00%	3.39%	
146	LU	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.39% S	3.20%	Pass
146	LU	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.71%	0.99%	Pass
146	LU	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	11.41% S	5.39%	No pass
146	LU	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	9.87% S	7.15%	Pass
153	SI	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.47%	0.99%	Pass
153	SI	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	11.73% S	10.43%	Pass
153	SI	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-14.75% S	4.23%	Pass
153	SI	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.73% S	4.87%	Pass
153	SI	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.10% S	3.16%	Pass
153	SI	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.05% S	1.53%	Pass
153	SI	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-17.33% S	11.31%	
153	SI	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.37% S	1.44%	Pass
153	SI	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-10.93% S	5.19%	Pass
153	SI	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.85% S	0.50%	Pass
155	UK	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.21%	1.65%	Pass
155	UK	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-11.12% S	2.87%	Pass
155	UK	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.55% S	2.19%	Pass
155	UK	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.90% S	0.28%	Pass
155	UK	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.77% S	0.86%	Pass
155	UK	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.57% S	0.45%	Pass
155	UK	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.74% S	1.31%	
155	UK	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.98% S	2.88%	Pass
155	UK	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.39% S	0.99%	Pass
155	UK	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.48%	0.62%	Pass
158	ASIA	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.30% S	1.16%	Pass
158	ASIA	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.18% S	1.04%	Pass
158	ASIA	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.24% S	3.50%	Pass
158	ASIA	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.82% S	2.39%	Pass
158	ASIA	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.98%	6.57%	Pass
158	ASIA	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.62% S	2.01%	Pass
158	ASIA	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-14.88%	11.69%	No pass
158	ASIA	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.27% S	3.43%	Pass
158	ASIA	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-16.80% S	5.56%	
158	ASIA	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.07% S	1.44%	Pass
158	ASIA	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.68%	7.36%	Pass
158	ASIA	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.15% S	0.88%	Pass
158	ASIA	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.04% S	2.43%	Pass
165	VN	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.68%	5.45%	Pass

165	VN	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-21.67% S	10.55%	Pass
165	VN	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.19%	23.51%	Pass
165	VN	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	40.85% S	15.65%	No pass
165	VN	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-16.55% S	12.55%	No pass
165	VN	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	102.68% S	170.15%	No pass
165	VN	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.06% S	3.54%	Pass
165	VN	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	127.78% S	42.79%	No pass
165	VN	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.49%	3.86%	Pass
165	VN	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	54.65% S	11.26%	Pass
165	VN	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	4.98% S	2.87%	Pass
165	VN	NH3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.41%	7.03%	Pass
165	VN	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	190.28% S	78.51%	No pass
165	VN	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.64%	3.16%	Pass
165	VN	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.80% S	4.42%	
165	VN	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.31%	8.16%	Pass
165	VN	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.18%	2.59%	Pass
165	VN	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.48%	4.75%	Pass
165	VN	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.59% S	4.27%	Pass
165	VN	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	45.67%	13.76%	Pass
166	PL	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.61%	0.99%	Pass
166	PL	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
166	PL	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
166	PL	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.53% S	2.87%	Pass
166	PL	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.67%	3.36%	Pass
166	PL	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
166	PL	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.24%	2.09%	Pass
166	PL	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.42%	0.29%	Pass
166	PL	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.49%	1.09%	Pass
166	PL	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.32%	0.86%	Pass
166	PL	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
166	PL	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.91% S	1.80%	Pass
166	PL	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-7.74% S	5.21%	
166	PL	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.44%	3.68%	Pass
166	PL	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.47%	1.22%	Pass
166	PL	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.06% S	2.54%	Pass
166	PL	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.37%	0.32%	Pass
167	UK	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.14% S	0.99%	Pass
167	UK	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.39% S	4.17%	Pass
167	UK	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	17.98% S	1.75%	Pass
167	UK	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
167	UK	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.77% S	2.01%	Pass

167	UK	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.35% S	0.81%	Pass
167	UK	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
167	UK	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	31.91% S	14.89%	Pass
167	UK	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.53%	1.33%	Pass
167	UK	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.59% S	2.31%	Pass
171	FR	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.01% S	5.19%	Pass
171	FR	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	4.17% S	3.22%	Pass
171	FR	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	0.95%	5.32%	Pass
171	FR	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.28% S	3.29%	Pass
171	FR	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.21% S	4.19%	Pass
171	FR	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	6.34% S	4.22%	Pass
171	FR	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	14.36% S	11.28%	Pass
173	HR	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.24% S	1.17%	Pass
173	HR	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-24.93%	18.04%	No pass
174	RU	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
174	RU	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.43% S	10.27%	Pass
174	RU	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
174	RU	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-16.11% S	19.04%	Pass
174	RU	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	55.00% S	39.04%	No pass
174	RU	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
174	RU	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
176	RU	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.93%	3.14%	Pass
176	RU	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.86% S	6.52%	Pass
176	RU	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.29%	14.74%	No pass
176	RU	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-14.10% S	4.41%	Pass
176	RU	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	9.12% S	2.87%	Pass
176	RU	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.57% S	1.53%	Pass
176	RU	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	3.52% S	2.18%	
176	RU	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.82% S	1.92%	Pass
176	RU	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	12.95% S	5.86%	Pass
176	RU	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.27% S	8.51%	No pass
178	RU	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	17.40%	41.61%	No pass
178	RU	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.88% S	2.34%	Pass
178	RU	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-8.33%	1.24%	Pass
178	RU	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.11%	4.70%	Pass
178	RU	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.86%	137.91%	No pass
178	RU	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-13.91% S	8.22%	Pass
178	RU	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.31% S	2.36%	Pass
178	RU	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.37% S	5.85%	Pass
178	RU	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.68% S	4.92%	Pass
178	RU	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.08% S	0.57%	Pass

178	RU	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-8.94% S	10.36%	Pass
178	RU	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
178	RU	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-24.06% S	16.18%	
178	RU	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-33.82% S	7.68%	Pass
178	RU	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-20.30% S	6.41%	Pass
178	RU	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	182.41% S	64.63%	No pass
178	RU	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-13.42% S	6.66%	Pass
179	RU	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	11.98% S	580.55%	No pass
179	RU	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-22.61% S	4.51%	Pass
179	RU	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-23.27% S	9.15%	Pass
179	RU	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-11.99% S	14.61%	Pass
179	RU	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-19.40% S	6.86%	Pass
179	RU	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.77%	11.38%	Pass
179	RU	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
179	RU	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-5.88%	12.55%	Pass
179	RU	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.83%	28.38%	No pass
179	RU	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-31.61% S	2.87%	Pass
179	RU	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	7.83%	7.54%	Pass
179	RU	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-24.64% S	7.66%	No pass
179	RU	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
179	RU	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	19.39% S	12.65%	Pass
179	RU	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.54%	4.31%	Pass
179	RU	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-22.61% S	43.77%	No pass
179	RU	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.33%	4.45%	Pass
181	NL	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.96% S	1.32%	Pass
181	NL	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.88% S	0.92%	Pass
181	NL	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-9.33%	2.47%	Pass
181	NL	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	12.56% S	7.56%	Pass
181	NL	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	10.42% S	5.40%	Pass
181	NL	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00%	2.11%	Pass
181	NL	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.26% S	1.25%	Pass
181	NL	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.97% S	0.71%	Pass
181	NL	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00% S	0.00%	Pass
181	NL	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.80% S	0.86%	Pass
181	NL	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.33% S	2.09%	Pass
181	NL	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.40%	2.25%	Pass
181	NL	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-23.26% S	8.38%	
181	NL	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	5.78% S	4.16%	Pass
181	NL	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	4.42% S	1.66%	Pass
181	NL	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.59% S	0.58%	Pass
181	NL	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.39%	1.76%	Pass

182	RO	HNO3-N on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-11.36% S	4.11%	Pass
182	RO	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.45% S	0.78%	Pass
182	RO	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	435.04%	188.69%	No pass
182	RO	SO2-S on impregnated filter	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-17.85%	11.95%	No pass
183	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.94% S	0.78%	Pass
183	RS	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.82%	5.55%	Pass
184	RS	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-10.00% S	0.59%	Pass
184	RS	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	3.59%	0.87%	Pass
184	RS	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-4.06% S	0.84%	Pass
184	RS	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
184	RS	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.33%	3.49%	Pass
184	RS	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
184	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.00%	3.50%	Pass
184	RS	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	8.13%	1.80%	Pass
184	RS	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.39%	0.71%	Pass
185	ORG	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-1.90% S	0.78%	Pass
186	RS	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-13.15% S	10.94%	Pass
186	RS	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
186	RS	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-6.97%	6.34%	Pass
186	RS	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
186	RS	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
186	RS	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
186	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-0.80%	1.56%	Pass
186	RS	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	8.02% S	4.16%	Pass
186	RS	Zinc	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
187	COM	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.90%	7.10%	No pass
187	COM	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	1.77%	14.24%	Pass
187	COM	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	123.81% S	11.13%	Pass
187	COM	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-3.63%	14.35%	Pass
187	COM	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
187	COM	Chromium	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	6.97%	11.06%	Pass
187	COM	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf			
187	COM	Copper	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-52.78% S	23.89%	Pass
187	COM	Lead	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.07%	7.62%	Pass
187	COM	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-11.74% S	0.86%	Pass
187	COM	Nickel	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	40.00%	45.01%	No pass
187	COM	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.88% S	1.53%	Pass
187	COM	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-2.25% S	1.17%	Pass
187	COM	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	-16.80%	14.33%	
187	COM	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	0.25%	24.98%	No pass
187	COM	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccq/qameasure/emep31.pdf	2.01%	6.52%	Pass

187	COM	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-17.78%	11.93%	No pass
187	COM	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-42.82% S	7.39%	No pass
187	COM	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-9.52% S	2.82%	Pass
188	COM	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.11% S	1.17%	Pass
188	COM	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	7.40%	2.36%	Pass
189	COM	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-6.67%	12.56%	Pass
189	COM	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
189	COM	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.18% S	0.00%	Pass
189	COM	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
189	COM	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-26.84% S	1.88%	Pass
189	COM	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.67% S	2.09%	Pass
189	COM	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.43%	3.50%	Pass
189	COM	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.72%	3.19%	Pass
189	COM	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.30%	3.46%	Pass
190	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.43%	1.56%	Pass
190	RS	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-44.37%	25.94%	No pass
192	COM	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-0.86%	3.11%	Pass
192	COM	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-8.13% S	12.07%	No pass
193	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.75% S	2.33%	Pass
193	RS	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.90%	34.27%	No pass
194	RS	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.92%	2.51%	Pass
194	RS	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-7.61% S	11.45%	No pass
194	RS	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	5.48% S	2.30%	Pass
194	RS	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.68% S	0.78%	Pass
194	RS	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	19.05%	36.62%	
194	RS	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.14%	1.94%	Pass
194	RS	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
194	RS	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-1.86% S	0.00%	Pass
195	RS	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.65%	10.73%	No pass
195	RS	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-10.82% S	2.34%	Pass

195	RS	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.57%	7.42%	Pass
195	RS	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-84.53% S	11.74%	Pass
195	RS	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-87.41% S	33.86%	No pass
195	RS	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-45.48% S	14.17%	Pass
195	RS	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
195	RS	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-40.56% S	9.09%	Pass
195	RS	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.42% S	3.79%	Pass
195	RS	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-20.53% S	3.16%	Pass
195	RS	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-39.03% S	8.37%	Pass
195	RS	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	1.20%	2.71%	Pass
195	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.44% S	0.39%	Pass
195	RS	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	8.40% S	5.24%	
195	RS	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-15.85%	12.49%	Pass
195	RS	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-23.30% S	8.62%	Pass
195	RS	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
195	RS	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-3.64%	8.60%	Pass
196	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.76% S	1.94%	Pass
197	COM	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	16.85%	16.02%	No pass
197	COM	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-40.43% S	1.24%	Pass
197	COM	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-44.44% S	7.13%	Pass
197	COM	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-29.43% S	0.00%	Pass
197	COM	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	424.36% S	157.08%	No pass
197	COM	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-84.77% S	23.33%	No pass
197	COM	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
197	COM	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	173.90% S	65.60%	No pass
197	COM	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-16.28% S	0.71%	Pass
198	RS	Ammonium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-4.46%	5.12%	Pass
198	RS	Arsenic	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Cadmium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Calcium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Chloride in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Chromium	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-64.10% S	10.54%	Pass

198	RS	Conductivity in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Copper	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Lead	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Magnesium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	26.51% S	8.91%	Pass
198	RS	Nickel	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	10.00%	15.70%	Pass
198	RS	Nitrate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	NO2-N in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-2.25%	1.94%	Pass
198	RS	pH in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Potassium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-8.17% S	2.40%	Pass
198	RS	SO2-S in absorbing solution	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-5.00%	3.33%	Pass
198	RS	Sodium in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	-53.24% S	3.92%	Pass
198	RS	Sulphate in precipitation	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf			
198	RS	Zinc	EMEP31	20131201	https://projects.nilu.no/ccc/qameasure/emep31.pdf	2.18%	2.12%	Pass