

Lab	Lab name	Component	QA measure ID	QA date	QA document url	QA bias	QA variability	QA outcome
3	CZ	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.94% S	2.42%	Pass
3	CZ	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.67%	14.28%	Pass
3	CZ	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.36%	9.33%	Pass
3	CZ	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	24.08% S	4.26%	Pass
3	CZ	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.42% S	0.51%	Pass
3	CZ	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-18.50% S	4.21%	Pass
3	CZ	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.97% S	1.89%	Pass
3	CZ	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.85% S	0.46%	Pass
3	CZ	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.63%	4.94%	Pass
3	CZ	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-10.91% S	2.66%	Pass
3	CZ	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.03% S	1.90%	Pass
3	CZ	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.60%	4.37%	Pass
3	CZ	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.83% S	1.84%	Pass
3	CZ	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.89%	1.76%	Pass
3	CZ	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	6.37% S	4.78%	
3	CZ	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.15%	2.71%	Pass
3	CZ	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	22.39%	16.84%	No pass
3	CZ	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.71% S	1.18%	Pass
3	CZ	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.11%	2.64%	Pass
3	CZ	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	21.84% S	4.39%	Pass
4	DK	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.79% S	0.38%	Pass
4	DK	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	20.42% S	3.27%	Pass
4	DK	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.57%	7.78%	Pass
4	DK	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-20.13% S	3.79%	Pass
4	DK	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.42%	1.43%	Pass
4	DK	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	26.67% S	8.59%	Pass
4	DK	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
4	DK	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	20.12% S	3.35%	Pass
4	DK	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-12.88% S	6.03%	Pass
4	DK	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.78% S	1.51%	Pass
4	DK	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-18.04% S	3.48%	Pass
4	DK	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-9.42% S	2.13%	Pass
4	DK	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	6.56% S	1.02%	Pass
4	DK	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.56%	0.59%	Pass
4	DK	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.74%	4.13%	
4	DK	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-19.52% S	3.34%	Pass
4	DK	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-21.21% S	4.51%	Pass
4	DK	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.70% S	1.58%	Pass
4	DK	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.88% S	0.41%	Pass
4	DK	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			

5	FI	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.92% S	0.13%	Pass
5	FI	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.08% S	0.87%	Pass
5	FI	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.71% S	0.78%	Pass
5	FI	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.26% S	0.24%	Pass
5	FI	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.13%	0.51%	Pass
5	FI	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.79%	0.86%	Pass
5	FI	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.46%	1.30%	Pass
5	FI	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.87% S	0.50%	Pass
5	FI	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.78% S	0.87%	Pass
5	FI	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.13%	0.48%	Pass
5	FI	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.82% S	1.27%	Pass
5	FI	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.66% S	2.10%	Pass
5	FI	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.39%	1.29%	Pass
5	FI	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.15%	0.49%	Pass
5	FI	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	28.86% S	10.68%	
5	FI	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.79% S	1.04%	Pass
5	FI	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	14.48% S	2.58%	Pass
5	FI	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.35%	1.64%	Pass
5	FI	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.16%	1.02%	Pass
5	FI	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.89% S	1.07%	Pass
6	COM	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.28%	9.64%	No pass
6	COM	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	10.00%	2.18%	Pass
6	COM	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	9.52%	7.78%	Pass
6	COM	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.57%	2.18%	Pass
6	COM	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.81% S	0.77%	Pass
6	COM	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-50.00% S	0.00%	Pass
6	COM	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.03% S	1.24%	Pass
6	COM	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.29%	0.84%	Pass
6	COM	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.69% S	2.10%	Pass
6	COM	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-12.27% S	0.60%	Pass
6	COM	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.23%	0.66%	Pass
6	COM	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	14.69% S	22.01%	No pass
6	COM	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-20.57% S	1.36%	Pass
6	COM	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.23% S	0.57%	Pass
6	COM	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	53.36%	59.43%	
6	COM	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.69% S	0.27%	Pass
6	COM	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-11.25% S	7.26%	Pass
6	COM	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.39%	1.59%	Pass
6	COM	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.84% S	1.84%	Pass
6	COM	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	42.22% S	5.06%	Pass
7	COM	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.19% S	0.38%	Pass

7	COM	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
7	COM	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00%	3.89%	Pass
7	COM	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	4.96% S	1.42%	Pass
7	COM	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.39% S	0.78%	Pass
7	COM	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
7	COM	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-6.72% S	2.15%	Pass
7	COM	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.03% S	0.84%	Pass
7	COM	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.22%	0.60%	Pass
7	COM	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.89%	1.58%	Pass
7	COM	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.68% S	0.00%	Pass
7	COM	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.73%	0.29%	Pass
7	COM	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-18.52%	54.79%	
7	COM	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	10.17% S	4.59%	Pass
7	COM	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-14.01% S	7.70%	Pass
7	COM	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.77% S	1.02%	Pass
7	COM	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	6.08% S	1.25%	Pass
8	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.68%	0.51%	Pass
8	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.80% S	2.98%	Pass
8	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.96% S	1.32%	Pass
8	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.43% S	0.71%	Pass
8	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.13%	1.52%	Pass
8	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.33% S	1.89%	Pass
8	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	7.67% S	1.90%	Pass
8	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.65% S	0.54%	Pass
8	DE	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.33%	2.72%	Pass
8	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.40% S	0.67%	Pass
8	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.32% S	0.95%	Pass
8	DE	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	5.57% S	1.45%	Pass
8	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.54% S	1.36%	Pass
8	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.43%	2.83%	Pass
8	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.33%	6.46%	
8	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.77%	4.59%	Pass
8	DE	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.49% S	0.24%	Pass
8	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.30% S	2.57%	Pass
8	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.82% S	1.02%	Pass
8	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.07%	0.65%	Pass
10	HU	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.24%	4.84%	Pass
10	HU	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
10	HU	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-35.71% S	4.67%	Pass
10	HU	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.78%	3.08%	Pass
10	HU	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.68%	2.48%	Pass

10	HU	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
10	HU	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.42% S	1.24%	Pass
10	HU	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
10	HU	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.55%	8.19%	Pass
10	HU	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.47%	5.87%	Pass
10	HU	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.36% S	1.27%	Pass
10	HU	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.12% S	3.47%	Pass
10	HU	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
10	HU	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.60% S	1.66%	Pass
10	HU	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.40% S	0.51%	Pass
10	HU	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	14.94% S	8.77%	
10	HU	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.55% S	0.84%	Pass
10	HU	SO2-S in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.06% S	0.84%	Pass
10	HU	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	20.55% S	2.44%	Pass
10	HU	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.92% S	0.59%	Pass
10	HU	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.88% S	1.32%	Pass
10	HU	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
12	IE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.20%	4.58%	Pass
12	IE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-13.25% S	2.13%	Pass
12	IE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.23% S	2.02%	Pass
12	IE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.19%	0.64%	Pass
12	IE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.49%	0.63%	Pass
12	IE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.42% S	0.98%	Pass
12	IE	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.69% S	1.70%	Pass
12	IE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-8.77% S	6.81%	
12	IE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.62% S	2.09%	Pass
12	IE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.27% S	2.83%	Pass
12	IE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.22% S	1.63%	Pass
15	NO	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.90% S	3.44%	Pass
15	NO	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00%	2.18%	Pass
15	NO	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.56% S	3.89%	Pass
15	NO	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.23% S	1.42%	Pass
15	NO	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.52%	1.79%	Pass
15	NO	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00% S	0.00%	Pass
15	NO	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.25% S	0.53%	Pass
15	NO	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00% S	0.00%	Pass
15	NO	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.50%	10.46%	No pass
15	NO	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00% S	0.00%	Pass
15	NO	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.91%	3.80%	Pass
15	NO	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.73%	2.57%	Pass
15	NO	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00% S	0.00%	Pass

15	NO	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.30% S	2.34%	Pass
15	NO	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-15.91% S	8.32%	Pass
15	NO	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	7.18% S	3.16%	
15	NO	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.31%	5.01%	Pass
15	NO	SO2-S in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.48%	1.12%	Pass
15	NO	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-8.92% S	5.07%	Pass
15	NO	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	4.67% S	0.53%	Pass
15	NO	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.15%	1.52%	Pass
15	NO	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	3.47% S	2.18%	Pass
16	COM	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.59%	2.42%	Pass
16	COM	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	12.91% S	6.63%	Pass
16	COM	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.63% S	2.16%	Pass
16	COM	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.08%	0.96%	Pass
16	COM	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.58%	13.27%	No pass
16	COM	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.69%	0.95%	Pass
16	COM	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.75%	1.93%	Pass
16	COM	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.46% S	2.44%	Pass
16	COM	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	9.65% S	4.18%	
16	COM	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.94% S	1.46%	Pass
16	COM	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	5.25%	5.35%	Pass
16	COM	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.34%	3.95%	Pass
16	COM	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.10% S	1.73%	Pass
16	PL	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
16	PL	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00% S	0.00%	Pass
16	PL	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-6.67% S	1.72%	Pass
16	PL	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-9.55% S	5.44%	Pass
16	PL	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00%	0.60%	Pass
16	PL	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-14.14% S	2.72%	Pass
16	PL	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.16% S	2.38%	Pass
16	PL	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00% S	0.00%	Pass
19	ES	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	25.19% S	6.49%	Pass
19	ES	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.50%	2.60%	Pass
19	ES	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.12%	6.63%	Pass
19	ES	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	6.33% S	2.31%	Pass
19	ES	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.31%	2.53%	Pass
19	ES	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	9.02%	6.03%	Pass
19	ES	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.35%	0.98%	Pass
19	ES	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.47% S	1.70%	Pass
19	ES	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	18.86% S	8.05%	
19	ES	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.37%	0.84%	Pass
19	ES	SO2-S in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-28.79% S	4.21%	Pass

19	ES	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.69% S	2.83%	Pass
19	ES	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.68% S	1.52%	Pass
20	SE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	14.13% S	1.65%	Pass
20	SE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	8.08% S	1.52%	Pass
20	SE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.94% S	0.78%	Pass
20	SE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00%	0.71%	Pass
20	SE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.99%	32.85%	No pass
20	SE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.83% S	0.69%	Pass
20	SE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.16% S	1.05%	Pass
20	SE	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.00% S	0.54%	Pass
20	SE	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.09%	0.71%	Pass
20	SE	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.79%	0.60%	Pass
20	SE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.91%	0.63%	Pass
20	SE	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.50%	2.51%	Pass
20	SE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.81% S	0.68%	Pass
20	SE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.28%	18.17%	No pass
20	SE	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.39%	2.55%	Pass
20	SE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-10.54% S	3.24%	
20	SE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.76%	3.76%	Pass
20	SE	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.99% S	3.57%	Pass
20	SE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.94% S	6.58%	Pass
20	SE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.80%	13.11%	No pass
20	SE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.61%	2.02%	Pass
23	COM	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.43%	1.53%	Pass
23	COM	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.11% S	4.73%	Pass
23	COM	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.38% S	1.33%	Pass
23	COM	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	8.41% S	10.89%	No pass
23	COM	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.39%	5.00%	Pass
23	COM	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.48% S	0.20%	Pass
23	COM	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.58% S	0.34%	Pass
23	COM	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	16.15% S	2.51%	
23	COM	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-15.59% S	4.05%	Pass
23	COM	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.61% S	1.25%	Pass
23	COM	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-8.42% S	1.02%	Pass
24	RS	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.02%	4.33%	Pass
24	RS	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
24	RS	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00%	0.00%	Pass
24	RS	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.41%	6.15%	Pass
24	RS	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
24	RS	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
24	RS	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.92%	0.22%	Pass

24	RS	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
24	RS	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
24	RS	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.94%	4.75%	Pass
24	RS	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
24	RS	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
24	RS	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	9.91% S	10.06%	
24	RS	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.02%	2.71%	Pass
24	RS	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.67% S	1.64%	Pass
24	RS	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
24	RS	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
26	CA	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.27% S	0.76%	Pass
26	CA	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.11% S	0.47%	Pass
26	CA	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.69%	1.24%	Pass
26	CA	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
26	CA	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.77% S	0.95%	Pass
26	CA	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.44%	1.27%	Pass
26	CA	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-9.84%	6.07%	
26	CA	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.60% S	1.67%	Pass
26	CA	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.11%	1.51%	Pass
26	CA	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.56% S	0.20%	Pass
27	EDU	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.10% S	3.95%	Pass
27	EDU	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.21% S	0.95%	Pass
27	EDU	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.87%	0.97%	Pass
27	EDU	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.18% S	1.27%	Pass
27	EDU	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.13% S	0.95%	Pass
27	EDU	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.95%	2.25%	Pass
27	EDU	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	28.82% S	8.83%	
27	EDU	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.80% S	1.25%	Pass
27	EDU	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.22%	1.78%	Pass
27	EDU	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.81% S	0.41%	Pass
30	EU	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.09%	1.27%	Pass
30	EU	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	9.50% S	2.13%	Pass
30	EU	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	7.79% S	7.82%	Pass
30	EU	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-13.21% S	4.92%	Pass
30	EU	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.12% S	1.58%	Pass
30	EU	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.04% S	0.78%	Pass
30	EU	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	28.96%	19.16%	
30	EU	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.92%	2.09%	Pass
30	EU	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.12%	1.12%	Pass
30	EU	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.11%	1.12%	Pass
31	SK	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.14%	1.40%	Pass

31	SK	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.33%	3.27%	Pass
31	SK	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-44.72% S	3.11%	Pass
31	SK	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-10.68% S	4.02%	Pass
31	SK	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.48% S	6.72%	Pass
31	SK	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-11.46% S	13.92%	Pass
31	SK	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	7.45% S	1.89%	Pass
31	SK	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-18.86% S	3.48%	Pass
31	SK	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.38%	7.37%	Pass
31	SK	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-9.51% S	4.90%	Pass
31	SK	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-9.16% S	2.22%	Pass
31	SK	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	8.52% S	0.68%	Pass
31	SK	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-18.47% S	4.97%	Pass
31	SK	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.42%	4.40%	Pass
31	SK	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.65% S	0.51%	Pass
31	SK	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-16.74% S	9.84%	
31	SK	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-6.13% S	2.09%	Pass
31	SK	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	38.64% S	11.86%	No pass
31	SK	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.49% S	1.97%	Pass
31	SK	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	17.03%	9.65%	No pass
31	SK	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-15.73% S	5.91%	Pass
32	LT	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	9.91% S	0.51%	Pass
32	LT	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
32	LT	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
32	LT	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.02%	8.99%	Pass
32	LT	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.27% S	1.29%	Pass
32	LT	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
32	LT	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.95%	3.14%	Pass
32	LT	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
32	LT	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-10.44% S	7.33%	Pass
32	LT	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
32	LT	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
32	LT	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.49%	4.08%	Pass
32	LT	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
32	LT	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.52%	1.07%	Pass
32	LT	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.87% S	1.87%	Pass
32	LT	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-8.41%	10.89%	
32	LT	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.29%	3.13%	Pass
32	LT	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	29.16% S	2.99%	Pass
32	LT	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.61% S	3.16%	Pass
32	LT	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.60%	1.52%	Pass
32	LT	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			

33	LV	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-24.04% S	5.22%	Pass
33	LV	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-36.67% S	6.53%	Pass
33	LV	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	25.00% S	3.89%	Pass
33	LV	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	7.16% S	4.26%	Pass
33	LV	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.92% S	2.85%	Pass
33	LV	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-8.33%	6.88%	Pass
33	LV	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.16% S	1.81%	Pass
33	LV	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-15.48% S	6.28%	Pass
33	LV	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.84% S	0.81%	Pass
33	LV	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	27.78% S	12.10%	Pass
33	LV	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00%	1.27%	Pass
33	LV	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.81%	2.76%	Pass
33	LV	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
33	LV	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.04% S	3.03%	Pass
33	LV	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.90%	1.02%	Pass
33	LV	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-13.89% S	5.11%	
33	LV	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.47%	1.88%	Pass
33	LV	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.79%	4.30%	Pass
33	LV	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	7.96%	10.00%	No pass
33	LV	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.90% S	5.69%	No pass
33	LV	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.12% S	2.72%	Pass
34	TR	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-63.63% S	19.60%	No pass
34	TR	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.00%	10.23%	Pass
34	TR	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-11.27%	12.05%	Pass
34	TR	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-25.90% S	1.66%	Pass
34	TR	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-22.33% S	5.71%	Pass
34	TR	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-9.42%	11.77%	Pass
34	TR	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	12.03% S	10.35%	No pass
34	TR	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	8.33% S	3.48%	Pass
34	TR	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-99.12% S	50.70%	No pass
34	TR	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-11.14% S	4.17%	Pass
34	TR	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-14.54% S	3.80%	Pass
34	TR	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-96.33% S	35.93%	No pass
34	TR	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	415.45%	494.01%	No pass
34	TR	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-83.64% S	39.56%	No pass
34	TR	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-74.48% S	28.01%	No pass
34	TR	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-68.89% S	37.98%	
34	TR	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.99%	5.01%	Pass
34	TR	SO2-S in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-75.07% S	25.41%	No pass
34	TR	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-98.23% S	50.68%	No pass
34	TR	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.45%	9.54%	Pass

34	TR	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-75.75% S	20.93%	No pass
34	TR	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-13.20% S	4.12%	Pass
35	HR	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.98% S	2.93%	Pass
35	HR	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.94% S	1.18%	Pass
35	HR	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.21% S	0.37%	Pass
35	HR	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.02%	1.04%	Pass
35	HR	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.26% S	0.63%	Pass
35	HR	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.93% S	0.49%	Pass
35	HR	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	9.65% S	4.11%	
35	HR	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.22%	0.84%	Pass
35	HR	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.17%	0.99%	Pass
35	HR	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.52% S	0.91%	Pass
36	SI	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.05%	9.42%	No pass
36	SI	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-50.00% S	0.00%	Pass
36	SI	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	4.29%	0.00%	Pass
36	SI	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.20%	1.66%	Pass
36	SI	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.21% S	3.27%	Pass
36	SI	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
36	SI	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.64%	1.16%	Pass
36	SI	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.45%	0.42%	Pass
36	SI	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	12.45% S	2.36%	Pass
36	SI	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.46% S	0.95%	Pass
36	SI	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-50.00% S	0.00%	Pass
36	SI	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.38%	0.49%	Pass
36	SI	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.09%	4.03%	
36	SI	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.94% S	1.88%	Pass
36	SI	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.30%	2.11%	Pass
36	SI	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	3.01% S	1.02%	Pass
36	SI	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.79% S	1.03%	Pass
38	EE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.65% S	1.40%	Pass
38	EE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-10.00% S	4.35%	Pass
38	EE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-33.33% S	3.89%	Pass
38	EE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.78%	3.08%	Pass
38	EE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.80% S	4.05%	Pass
38	EE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
38	EE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.97%	2.42%	Pass
38	EE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-56.44% S	4.19%	Pass
38	EE	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-11.61% S	6.12%	Pass
38	EE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-24.39% S	5.44%	Pass
38	EE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.85% S	0.95%	Pass
38	EE	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	9.36% S	2.63%	Pass

38	EE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-36.07% S	14.97%	Pass
38	EE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.48% S	0.88%	Pass
38	EE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-32.63% S	16.50%	
38	EE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.00% S	2.30%	Pass
38	EE	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-29.27% S	7.86%	Pass
38	EE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.44% S	2.17%	Pass
38	EE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.82% S	0.71%	Pass
38	EE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.90%	5.66%	Pass
39	PL	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.57%	2.29%	Pass
39	PL	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	41.67%	0.00%	Pass
39	PL	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.56%	3.89%	Pass
39	PL	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.61% S	0.24%	Pass
39	PL	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.89% S	2.21%	Pass
39	PL	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00% S	0.00%	Pass
39	PL	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.83%	1.67%	Pass
39	PL	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-10.10% S	0.00%	Pass
39	PL	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.04% S	18.86%	No pass
39	PL	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.55%	3.02%	Pass
39	PL	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.02%	1.90%	Pass
39	PL	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-11.06% S	4.14%	Pass
39	PL	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-17.14%	6.80%	Pass
39	PL	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.66% S	1.37%	Pass
39	PL	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.95% S	1.36%	Pass
39	PL	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	9.65% S	1.64%	
39	PL	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.47%	1.46%	Pass
39	PL	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.73%	10.49%	No pass
39	PL	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.50%	3.42%	Pass
39	PL	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.18%	1.93%	Pass
39	PL	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00% S	0.00%	Pass
40	MK	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
40	MK	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
40	MK	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
40	MK	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-6.27% S	2.00%	Pass
40	MK	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
40	MK	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
40	MK	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-27.38%	14.82%	
40	MK	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
40	MK	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
40	MK	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
41	NET	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-15.58%	14.81%	Pass
41	NET	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			

41	NET	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	152.50%	0.00%	Pass
41	NET	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.58% S	1.93%	Pass
41	NET	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.45%	75.48%	No pass
41	NET	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-50.00% S	13.61%	Pass
41	NET	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	17.81% S	14.32%	Pass
42	COM	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	27.38%	27.86%	No pass
42	COM	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	99.31% S	58.03%	No pass
42	COM	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	240.89% S	296.59%	No pass
42	COM	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	69.81%	85.98%	No pass
42	COM	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	18.30% S	12.88%	No pass
42	COM	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	16.85% S	6.91%	No pass
42	COM	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.81%	9.00%	Pass
42	COM	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-15.74% S	22.41%	
42	COM	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	535.73% S	1253.02%	No pass
42	COM	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	9.75%	11.13%	No pass
42	COM	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	33.56% S	7.46%	No pass
45	COM	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	88.17%	117.48%	No pass
45	COM	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.32%	1.89%	Pass
45	COM	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.16% S	27.26%	No pass
45	COM	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.29% S	0.64%	Pass
45	COM	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.65% S	1.27%	Pass
45	COM	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.59% S	0.66%	Pass
45	COM	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-38.88% S	22.21%	
45	COM	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.34%	2.92%	Pass
45	COM	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.07%	6.78%	Pass
45	COM	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.33% S	1.19%	Pass
46	PL	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
46	PL	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
46	PL	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
46	PL	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.47%	1.67%	Pass
46	PL	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
46	PL	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
46	PL	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	13.51% S	3.36%	
46	PL	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
46	PL	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
46	PL	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
48	BE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.17% S	2.18%	Pass
48	BE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.54%	6.22%	Pass
48	BE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-59.67% S	8.08%	Pass
48	BE	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-24.62% S	3.64%	Pass
48	BE	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.44% S	2.12%	Pass

48	BE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-29.45% S	3.61%	Pass
48	BE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.21%	2.59%	Pass
49	CY	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.14% S	3.18%	Pass
49	CY	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.15% S	3.31%	Pass
49	CY	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.70% S	0.41%	Pass
49	CY	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.74% S	1.37%	Pass
49	CY	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.97%	2.22%	Pass
49	CY	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.50%	1.07%	Pass
49	CY	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.85%	14.61%	
49	CY	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.59% S	6.47%	Pass
49	CY	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.56%	2.17%	Pass
49	CY	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.16%	1.02%	Pass
50	FR	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	7.36% S	1.27%	Pass
50	FR	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-16.31% S	4.73%	Pass
50	FR	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.57% S	1.38%	Pass
50	FR	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.64% S	2.13%	Pass
50	FR	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.05% S	2.22%	Pass
50	FR	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.03% S	1.47%	Pass
50	FR	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	40.22% S	11.31%	
50	FR	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.41%	3.97%	Pass
50	FR	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.66% S	3.09%	Pass
50	FR	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.83% S	0.71%	Pass
51	COM	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00%	0.00%	Pass
51	COM	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
51	COM	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
51	COM	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-8.71%	4.19%	Pass
51	COM	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.55%	3.02%	Pass
51	COM	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-14.29% S	10.21%	Pass
51	COM	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-16.11% S	2.18%	Pass
52	PL	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	8.33% S	2.18%	Pass
52	PL	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.78% S	1.56%	Pass
52	PL	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
52	PL	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
52	PL	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.69% S	0.60%	Pass
52	PL	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-13.71% S	1.36%	Pass
52	PL	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
110	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.75%	4.58%	Pass
110	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00%	6.53%	Pass
110	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.05%	3.89%	Pass
110	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.56%	1.18%	Pass
110	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.65%	2.12%	Pass

110	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	45.00% S	12.03%	Pass
110	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.52%	1.48%	Pass
110	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.61% S	1.67%	Pass
110	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.02%	1.21%	Pass
110	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.04% S	1.27%	Pass
110	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.18%	3.40%	Pass
110	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.99%	3.42%	Pass
110	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	3.52% S	1.54%	
110	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.99% S	9.19%	Pass
110	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.87%	3.03%	Pass
110	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.51%	1.22%	Pass
110	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.83% S	2.18%	Pass
112	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	4.82% S	3.31%	Pass
112	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
112	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	4.64% S	1.56%	Pass
112	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.17%	1.42%	Pass
112	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.40% S	2.16%	Pass
112	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.88% S	1.55%	Pass
112	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.30% S	2.09%	Pass
112	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.10% S	32.87%	No pass
112	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.03%	0.30%	Pass
112	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.31%	2.53%	Pass
112	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-12.76% S	0.27%	Pass
112	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.26% S	0.49%	Pass
112	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.33%	6.65%	
112	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.66% S	1.67%	Pass
112	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.54% S	0.86%	Pass
112	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.52%	0.91%	Pass
112	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	19.00% S	8.16%	Pass
114	IT	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.34%	2.55%	Pass
114	IT	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	400.00%	87.09%	No pass
114	IT	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
114	IT	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	35.03% S	17.99%	No pass
114	IT	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.38%	2.94%	Pass
114	IT	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00%	8.59%	Pass
114	IT	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.11%	1.05%	Pass
114	IT	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00% S	8.37%	Pass
114	IT	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	35.42% S	18.14%	Pass
114	IT	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.07% S	1.27%	Pass
114	IT	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-17.14% S	0.00%	Pass
114	IT	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.61%	2.44%	Pass

114	IT	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.48%	4.60%	
114	IT	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	5.46%	10.44%	Pass
114	IT	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.96% S	2.83%	Pass
114	IT	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.02% S	1.63%	Pass
114	IT	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-15.83% S	3.27%	Pass
115	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.07%	0.38%	Pass
115	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	12.67% S	7.40%	Pass
115	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	8.73% S	2.33%	Pass
115	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-7.32% S	3.55%	Pass
115	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.65% S	0.87%	Pass
115	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.63%	0.86%	Pass
115	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.32% S	2.72%	Pass
115	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.76%	1.67%	Pass
115	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.23% S	1.15%	Pass
115	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-6.38% S	1.58%	Pass
115	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.68%	2.65%	Pass
115	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.01% S	0.88%	Pass
115	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	13.51% S	0.71%	
115	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-17.13% S	14.62%	Pass
115	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.60%	3.62%	Pass
115	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.59% S	0.91%	Pass
115	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.18% S	3.27%	Pass
116	CH	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.49% S	1.53%	Pass
116	CH	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
116	CH	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-9.55% S	5.74%	Pass
116	CH	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-8.86% S	5.23%	Pass
116	CH	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
116	CH	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.80% S	5.57%	No pass
116	CH	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	9.68%	22.51%	
116	CH	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
116	CH	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
116	CH	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.23%	0.63%	Pass
117	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-7.61% S	1.65%	Pass
117	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
117	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-32.50%	27.99%	No pass
117	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.48% S	0.71%	Pass
117	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-12.29% S	6.17%	Pass
117	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	7.50%	0.00%	Pass
117	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.19%	3.75%	Pass
117	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-20.96% S	6.28%	Pass
117	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	28.99% S	22.98%	Pass

117	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.94% S	0.32%	Pass
117	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
117	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.03% S	2.34%	Pass
117	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	13.69% S	8.92%	
117	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-21.32% S	5.85%	Pass
117	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-17.80% S	8.49%	Pass
117	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.69% S	0.30%	Pass
117	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	22.29% S	7.62%	Pass
118	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	5.30% S	2.29%	Pass
118	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1583.33% S	130.64%	No pass
118	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-19.44% S	1.56%	Pass
118	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-12.10% S	3.79%	Pass
118	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.94% S	0.78%	Pass
118	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	683.33% S	77.35%	No pass
118	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-6.27% S	4.29%	Pass
118	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	20.20% S	16.75%	Pass
118	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00%	1.81%	Pass
118	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-7.43% S	1.58%	Pass
118	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
118	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	3.76% S	0.88%	Pass
118	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-10.85%	5.26%	
118	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-9.66% S	1.67%	Pass
118	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-11.72% S	5.46%	Pass
118	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.08%	1.73%	Pass
118	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	7.38% S	3.37%	Pass
120	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	4.82% S	2.93%	Pass
120	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	7.25% S	5.44%	Pass
120	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.06%	12.44%	Pass
120	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.58% S	0.71%	Pass
120	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.38%	4.42%	Pass
120	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	10.83% S	1.72%	Pass
120	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	3.26%	3.47%	Pass
120	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.33% S	1.51%	Pass
120	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.22%	1.15%	Pass
120	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.85% S	1.27%	Pass
120	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.25%	5.65%	Pass
120	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	3.01% S	1.66%	Pass
120	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	14.85%	9.24%	
120	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.54%	1.04%	Pass
120	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.35%	1.78%	Pass
120	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.82% S	1.42%	Pass

120	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.56%	3.58%	Pass
121	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.11%	2.16%	Pass
121	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-26.33% S	20.03%	Pass
121	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.10%	10.11%	Pass
121	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.19%	1.18%	Pass
121	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-6.82% S	2.94%	Pass
121	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.29%	1.20%	Pass
121	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	221.50% S	86.18%	No pass
121	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.56%	0.88%	Pass
121	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.48% S	1.15%	Pass
121	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.59% S	0.32%	Pass
121	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.48%	6.74%	Pass
121	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.39%	1.56%	Pass
121	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1023.21% S	387.04%	
121	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.21% S	2.09%	Pass
121	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.65% S	4.80%	Pass
121	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	19.39% S	2.64%	Pass
121	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.30% S	1.27%	Pass
125	DE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.96%	2.42%	Pass
125	DE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	23.00% S	15.46%	Pass
125	DE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.64%	1.17%	Pass
125	DE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.47% S	1.89%	Pass
125	DE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.52% S	1.84%	Pass
125	DE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.46%	3.78%	Pass
125	DE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	19.68% S	41.05%	No pass
125	DE	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-41.54% S	2.64%	Pass
125	DE	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.53% S	1.87%	Pass
125	DE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.41% S	2.06%	Pass
125	DE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.19% S	1.63%	Pass
125	DE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.26%	0.98%	Pass
125	DE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	10.11%	15.64%	
125	DE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.70% S	2.05%	Pass
125	DE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.64% S	3.49%	Pass
125	DE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.75% S	1.02%	Pass
125	DE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.02%	2.67%	Pass
126	IT	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-21.66% S	9.04%	No pass
126	IT	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	95.38% S	6.63%	Pass
126	IT	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.65%	4.28%	Pass
126	IT	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.14% S	3.68%	Pass
126	IT	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.37%	14.24%	No pass
126	IT	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.91%	2.64%	Pass

126	IT	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.23%	14.30%	
126	IT	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-18.53%	15.24%	Pass
126	IT	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-8.44%	8.75%	Pass
126	IT	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-8.28% S	2.74%	Pass
132	CL	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
132	CL	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
132	CL	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
132	CL	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
132	CL	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
132	CL	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
132	CL	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.67%	5.24%	Pass
141	JP	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
141	JP	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-10.56% S	4.67%	Pass
141	JP	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
141	JP	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
141	JP	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
141	JP	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
141	JP	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
145	EE	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.03%	6.49%	Pass
145	EE	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00%	1.52%	Pass
145	EE	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	8.27%	5.75%	Pass
145	EE	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	19.62% S	11.12%	Pass
145	EE	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.63% S	0.60%	Pass
145	EE	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
145	EE	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.77% S	0.33%	Pass
145	EE	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	10.00%	0.42%	Pass
145	EE	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.54%	1.27%	Pass
145	EE	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	10.00% S	3.48%	Pass
145	EE	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	11.44% S	3.88%	Pass
145	EE	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.80% S	0.59%	Pass
145	EE	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-15.61% S	13.39%	
145	EE	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-9.58% S	2.92%	Pass
145	EE	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.98% S	0.99%	Pass
145	EE	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.40% S	0.51%	Pass
145	EE	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.46% S	0.82%	Pass
146	LU	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-8.33% S	0.00%	Pass
146	LU	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.03%	4.67%	Pass
146	LU	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.96%	8.94%	Pass
146	LU	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.62%	1.63%	Pass
146	LU	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.11%	1.69%	Pass
146	LU	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.98%	2.18%	Pass

146	LU	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.49%	2.47%	Pass
153	SI	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.33%	4.20%	Pass
153	SI	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	12.89% S	8.52%	Pass
153	SI	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-8.61% S	4.23%	Pass
153	SI	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.60%	3.69%	Pass
153	SI	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.10%	6.01%	Pass
153	SI	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-6.18% S	2.54%	Pass
153	SI	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-9.69%	12.72%	
153	SI	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.26%	2.30%	Pass
153	SI	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	8.68% S	4.61%	Pass
153	SI	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.86% S	1.52%	Pass
155	UK	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.82%	1.27%	Pass
155	UK	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.50%	2.13%	Pass
155	UK	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.14% S	0.97%	Pass
155	UK	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.01%	1.71%	Pass
155	UK	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.77% S	0.63%	Pass
155	UK	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.93%	1.27%	Pass
155	UK	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-21.47% S	3.80%	
155	UK	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.90% S	1.88%	Pass
155	UK	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.62%	1.97%	Pass
155	UK	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.79% S	1.42%	Pass
158	ASIA	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.59% S	0.51%	Pass
158	ASIA	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.83% S	0.71%	Pass
158	ASIA	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.60%	2.90%	Pass
158	ASIA	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.46%	3.05%	Pass
158	ASIA	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.75%	1.36%	Pass
158	ASIA	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.04% S	0.32%	Pass
158	ASIA	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.53%	3.48%	Pass
158	ASIA	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.38%	2.05%	Pass
158	ASIA	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-37.54% S	18.01%	
158	ASIA	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.86% S	0.42%	Pass
158	ASIA	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-18.19% S	4.53%	Pass
158	ASIA	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.32%	1.45%	Pass
158	ASIA	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.03% S	0.41%	Pass
166	PL	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.97%	8.28%	Pass
166	PL	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.55% S	0.37%	Pass
166	PL	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			

166	PL	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	15.55% S	6.96%	Pass
166	PL	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.63%	0.68%	Pass
166	PL	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.11%	10.23%	Pass
166	PL	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
166	PL	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.76% S	0.51%	Pass
166	PL	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	9.17% S	3.27%	Pass
169	UK	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	32.08% S	7.62%	Pass
169	UK	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.74% S	2.80%	Pass
169	UK	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	23.33%	11.52%	Pass
169	UK	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.06% S	1.05%	Pass
169	UK	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.94% S	1.39%	Pass
169	UK	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-8.34% S	1.77%	Pass
169	UK	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.88% S	1.60%	Pass
171	FR	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.08%	1.09%	Pass
171	FR	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.78% S	0.78%	Pass
171	FR	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.33% S	1.72%	Pass
171	FR	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.16% S	1.09%	Pass
171	FR	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.94% S	1.21%	Pass
171	FR	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.31% S	0.75%	Pass
171	FR	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.36% S	3.88%	Pass
174	COM	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
174	COM	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-38.57% S	10.11%	Pass
174	COM	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
174	COM	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-41.46%	34.33%	No pass
174	COM	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-13.24% S	2.42%	Pass
174	COM	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
174	COM	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
178	RU	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
178	RU	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00% S	10.89%	Pass
178	RU	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	12.70% S	7.78%	Pass
178	RU	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.28% S	1.42%	Pass
178	RU	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
178	RU	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	166.67% S	8.59%	Pass
178	RU	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	5.02% S	2.12%	Pass
178	RU	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00% S	4.19%	Pass
178	RU	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.55% S	3.02%	Pass
178	RU	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.69%	2.85%	Pass

178	RU	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00% S	0.00%	Pass
178	RU	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
178	RU	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	18.86% S	4.28%	
178	RU	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	5.46%	4.18%	Pass
178	RU	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.45% S	2.83%	Pass
178	RU	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
178	RU	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	5.63% S	1.09%	Pass
179	RU	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	6.30%	4.45%	Pass
179	RU	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	58.33%	41.37%	No pass
179	RU	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.19%	10.11%	Pass
179	RU	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-17.74% S	6.15%	Pass
179	RU	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.91%	9.52%	Pass
179	RU	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.67%	18.91%	Pass
179	RU	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-5.70% S	1.24%	Pass
179	RU	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-11.69% S	2.09%	Pass
179	RU	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.91%	4.84%	Pass
179	RU	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	6.95% S	1.58%	Pass
179	RU	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	2.00%	4.76%	Pass
179	RU	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	1.26%	14.85%	No pass
179	RU	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf			
179	RU	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-24.68%	18.59%	Pass
179	RU	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-21.09% S	9.93%	Pass
179	RU	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.85%	2.13%	Pass
179	RU	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-12.75% S	3.27%	Pass
181	NL	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	4.21% S	1.65%	Pass
181	NL	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00%	4.35%	Pass
181	NL	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.00%	7.78%	Pass
181	NL	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.17%	1.42%	Pass
181	NL	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	0.52%	4.92%	Pass
181	NL	Chromium	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.25%	3.44%	Pass
181	NL	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.53%	9.13%	No pass
181	NL	Copper	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.47%	2.93%	Pass
181	NL	Lead	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.02% S	0.60%	Pass
181	NL	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.31%	2.53%	Pass
181	NL	Nickel	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-1.25%	6.12%	Pass
181	NL	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-4.91% S	2.93%	Pass
181	NL	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-12.88% S	2.87%	
181	NL	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-0.37%	0.84%	Pass
181	NL	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-2.22%	2.17%	Pass
181	NL	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	-3.02% S	2.54%	Pass
181	NL	Zinc	EMEP36	20181201	https://projects.nilu.no/ccq/qameasure/emep36.pdf	4.37%	5.44%	Pass

182	RO	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.30%	40.78%	No pass
182	RO	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.59%	13.01%	No pass
182	RO	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-36.56% S	18.16%	No pass
182	RO	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-91.95% S	53.98%	No pass
183	RS	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.73% S	1.36%	Pass
183	RS	SO2-S in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.21% S	1.26%	Pass
184	RS	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
184	RS	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
184	RS	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
184	RS	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
184	RS	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
184	RS	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
184	RS	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.28% S	1.19%	Pass
184	RS	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
185	ORG	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
185	ORG	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
185	ORG	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
185	ORG	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
185	ORG	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
185	ORG	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
185	ORG	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.06% S	0.34%	Pass
185	ORG	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
186	RS	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
186	RS	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-10.00%	34.99%	No pass
186	RS	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
186	RS	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
186	RS	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-9.83% S	3.75%	Pass
186	RS	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
186	RS	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.73% S	1.02%	Pass
186	RS	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-11.81% S	3.01%	Pass
187	COM	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
187	COM	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-14.00%	18.94%	Pass
187	COM	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.86%	25.66%	No pass
187	COM	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
187	COM	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
187	COM	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-13.08% S	11.86%	Pass
187	COM	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.43%	4.38%	Pass
187	COM	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.61%	5.61%	Pass
187	COM	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-10.02% S	3.63%	Pass
187	COM	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
187	COM	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-9.88%	6.74%	Pass

187	COM	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	23.86% S	149.24%	No pass
187	COM	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.11% S	0.34%	Pass
187	COM	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	217.57% S	130.09%	
187	COM	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
187	COM	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
187	COM	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
187	COM	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-11.78% S	3.92%	Pass
188	COM	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.57%	0.51%	Pass
189	RS	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
189	RS	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
189	RS	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
189	RS	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
189	RS	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
189	RS	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
189	RS	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.58% S	0.51%	Pass
189	RS	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
192	COM	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.41% S	1.02%	Pass
193	RS	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
193	RS	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
193	RS	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
193	RS	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.97%	1.48%	Pass
193	RS	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
193	RS	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
193	RS	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.47%	1.36%	Pass
193	RS	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-7.74%	7.20%	
193	RS	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
193	RS	SO2-S in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	3.75%	4.21%	Pass
193	RS	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
193	RS	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
194	RS	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
194	RS	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-17.87% S	1.42%	Pass
194	RS	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
194	RS	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
194	RS	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.81%	1.58%	Pass
194	RS	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
194	RS	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.67%	2.89%	Pass
194	RS	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
194	RS	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-6.09% S	2.09%	Pass
194	RS	SO2-S in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.41%	24.43%	No pass
194	RS	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.12% S	0.79%	Pass
194	RS	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			

197	COM	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.47%	1.53%	Pass
198	RS	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
198	RS	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
198	RS	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
198	RS	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
198	RS	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
198	RS	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
198	RS	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.01%	1.19%	Pass
198	RS	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
200	CH	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	4.19% S	3.31%	Pass
200	CH	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	2.00% S	1.42%	Pass
200	CH	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.04%	0.97%	Pass
200	CH	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	1.75%	1.15%	Pass
200	CH	HNO3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-0.96% S	2.90%	Pass
200	CH	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.00%	0.63%	Pass
200	CH	NH3-N on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-4.21% S	1.74%	Pass
200	CH	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.66%	0.59%	Pass
200	CH	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.73% S	0.34%	Pass
200	CH	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-11.59% S	4.86%	
200	CH	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	6.49% S	1.88%	Pass
200	CH	SO2-S on impregnated filter	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-2.45% S	1.09%	Pass
200	CH	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-1.43%	2.70%	Pass
200	CH	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	0.61% S	1.02%	Pass
203	COM	Arsenic	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	COM	Cadmium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	COM	Chromium	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	COM	Copper	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	COM	Lead	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	COM	Nickel	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	COM	NO2-N in absorbing solution	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	7.64% S	3.23%	Pass
203	COM	Zinc	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	RS	Ammonium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	RS	Calcium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	RS	Chloride in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			
203	RS	Conductivity in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-3.56%	3.91%	Pass
203	RS	Magnesium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-16.40%	357.30%	No pass
203	RS	Nitrate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	29.71% S	136.83%	No pass
203	RS	pH in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	126.01% S	113.30%	
203	RS	Potassium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-14.34%	220.73%	No pass
203	RS	Sodium in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf	-5.49%	606.62%	No pass
203	RS	Sulphate in precipitation	EMEP36	20181201	https://projects.nilu.no/ccc/qameasure/emep36.pdf			

206 COM

NO2-N in absorbing solution EMEP36

20181201 <https://projects.nilu.no/ccc/qameasure/emep36.pdf>

2.39% S

2.38%

Pass