

Lab	Lab name	Component	QA measure ID	QA date	QA document url	QA bias	QA variability	QA outcome
3	CZ	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.68% S	1.64%	Pass
3	CZ	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	23.92% S	4.14%	Pass
3	CZ	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	13.45% S	4.67%	Pass
3	CZ	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.22% S	4.96%	Pass
3	CZ	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.67% S	2.06%	Pass
3	CZ	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.17% S	2.92%	Pass
3	CZ	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.08%	2.67%	Pass
3	CZ	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.78% S	2.14%	Pass
3	CZ	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.46% S	1.03%	Pass
3	CZ	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.48% S	2.79%	Pass
3	CZ	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.76%	3.79%	Pass
3	CZ	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.21% S	0.95%	Pass
3	CZ	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.31% S	1.02%	Pass
3	CZ	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.66% S	5.47%	
3	CZ	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.77% S	2.40%	Pass
3	CZ	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.67% S	2.13%	Pass
3	CZ	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.07%	1.67%	Pass
3	CZ	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	31.67% S	4.53%	Pass
4	DK	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.12%	2.09%	Pass
4	DK	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.42%	5.44%	Pass
4	DK	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.33% S	1.58%	Pass
4	DK	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-17.48%	15.68%	No pass
4	DK	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.56%	2.89%	Pass
4	DK	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	17.33% S	2.23%	Pass
4	DK	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
4	DK	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	15.73% S	2.09%	Pass
4	DK	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.80%	6.70%	Pass
4	DK	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.89% S	2.42%	Pass
4	DK	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.35% S	10.56%	No pass
4	DK	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.04%	12.03%	No pass
4	DK	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-12.62% S	5.65%	Pass
4	DK	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.13%	2.45%	Pass
4	DK	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.15%	6.23%	
4	DK	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-20.68% S	3.33%	Pass
4	DK	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.01%	6.90%	Pass
4	DK	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.81%	1.44%	Pass
4	DK	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.57% S	0.39%	Pass
4	DK	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	21.71%	12.27%	Pass
5	FI	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.55%	1.49%	Pass

5	FI	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.58%	3.81%	Pass
5	FI	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.56%	4.18%	Pass
5	FI	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.14% S	0.52%	Pass
5	FI	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.48% S	0.92%	Pass
5	FI	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.45%	1.70%	Pass
5	FI	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.54%	2.23%	Pass
5	FI	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.65% S	0.41%	Pass
5	FI	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.97%	2.29%	Pass
5	FI	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.68% S	0.35%	Pass
5	FI	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.06% S	1.86%	Pass
5	FI	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.50%	2.91%	Pass
5	FI	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.37% S	1.98%	Pass
5	FI	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.03% S	0.82%	Pass
5	FI	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.52%	6.67%	
5	FI	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.83% S	2.40%	Pass
5	FI	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.23% S	2.27%	Pass
5	FI	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.42%	0.81%	Pass
5	FI	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.80% S	0.90%	Pass
5	FI	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.51% S	8.73%	Pass
6	COM	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.02%	8.05%	No pass
6	COM	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	21.67%	6.53%	Pass
6	COM	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	15.49% S	11.66%	Pass
6	COM	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.74% S	2.09%	Pass
6	COM	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.01% S	2.66%	Pass
6	COM	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.50%	0.00%	Pass
6	COM	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
6	COM	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.85% S	35.17%	No pass
6	COM	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.43%	1.56%	Pass
6	COM	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-13.81% S	1.21%	Pass
6	COM	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.30% S	3.42%	Pass
6	COM	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.16%	7.92%	Pass
6	COM	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.00%	19.05%	Pass
6	COM	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.90% S	3.78%	Pass
6	COM	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	137.91% S	30.83%	
6	COM	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.46%	0.74%	Pass
6	COM	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.74%	5.05%	Pass
6	COM	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.80% S	0.75%	Pass
6	COM	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.66% S	2.44%	Pass
6	COM	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	35.79% S	17.09%	Pass
7	COM	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	23.58% S	14.02%	No pass

7	COM	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
7	COM	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-25.00% S	6.98%		Pass
7	COM	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-11.11% S	1.05%		Pass
7	COM	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.42% S	1.74%		Pass
7	COM	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
7	COM	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.67%	1.68%		Pass
7	COM	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-23.47% S	9.42%		Pass
7	COM	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-53.96% S	6.05%		Pass
7	COM	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.71% S	0.62%		Pass
7	COM	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
7	COM	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	338.44% S	171.78%		No pass
7	COM	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-89.08% S	7.78%		
7	COM	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.67% S	1.85%		Pass
7	COM	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.42% S	5.13%		Pass
7	COM	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	194.86% S	54.39%		No pass
7	COM	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.25% S	2.29%		Pass
8	DE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.31%	0.60%		Pass
8	DE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.50% S	1.96%		Pass
8	DE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.26% S	3.90%		Pass
8	DE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.86% S	1.83%		Pass
8	DE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.42%	2.62%		Pass
8	DE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.00% S	1.89%		Pass
8	DE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.11% S	2.89%		Pass
8	DE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.77% S	0.96%		Pass
8	DE	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.53%	2.32%		Pass
8	DE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.10% S	2.06%		Pass
8	DE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.73%	1.24%		Pass
8	DE	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.53%	2.56%		Pass
8	DE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.99% S	2.72%		Pass
8	DE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.18%	2.45%		Pass
8	DE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.48%	5.70%		
8	DE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.48%	2.77%		Pass
8	DE	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.71%	2.11%		Pass
8	DE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.71%	1.25%		Pass
8	DE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.02%	2.06%		Pass
8	DE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.40%	1.39%		Pass
10	HU	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.51% S	2.09%		Pass
10	HU	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	316.00%	101.25%		No pass
10	HU	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-61.54% S	7.76%		Pass
10	HU	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.23%	16.72%		No pass

10	HU	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.68% S	10.00%	No pass
10	HU	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	214.83% S	4.64%	Pass
10	HU	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.58% S	2.01%	Pass
10	HU	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.53%	10.72%	Pass
10	HU	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-90.98%	54.54%	No pass
10	HU	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	124.79% S	23.59%	Pass
10	HU	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.25% S	3.42%	Pass
10	HU	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-97.50%	44.75%	No pass
10	HU	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	26.06%	21.50%	Pass
10	HU	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.12% S	3.07%	Pass
10	HU	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.63% S	1.74%	Pass
10	HU	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	17.77% S	9.02%	
10	HU	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	35.83% S	42.67%	No pass
10	HU	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-90.13%	54.46%	No pass
10	HU	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.38%	17.09%	No pass
10	HU	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.84% S	1.54%	Pass
10	HU	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.45%	6.37%	Pass
12	IE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.39%	15.81%	No pass
12	IE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-18.30% S	2.61%	Pass
12	IE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.24% S	2.43%	Pass
12	IE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.59%	1.06%	Pass
12	IE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.41% S	1.55%	Pass
12	IE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.03%	0.82%	Pass
12	IE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00%	1.89%	
12	IE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.79% S	2.22%	Pass
12	IE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.45% S	1.38%	Pass
12	IE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.36% S	1.03%	Pass
15	NO	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	9.06% S	4.92%	Pass
15	NO	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.17%	3.05%	Pass
15	NO	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.05% S	2.33%	Pass
15	NO	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.80%	2.35%	Pass
15	NO	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.16%	2.39%	Pass
15	NO	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.54%	1.89%	Pass
15	NO	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.62%	1.94%	Pass
15	NO	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.21%	2.09%	Pass
15	NO	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.38%	2.59%	Pass
15	NO	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.29%	1.39%	Pass
15	NO	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.31%	4.04%	Pass
15	NO	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.92%	6.12%	Pass
15	NO	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.25% S	1.36%	Pass

15	NO	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.56%	1.02%	Pass
15	NO	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.89% S	2.61%	Pass
15	NO	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.03%	5.14%	
15	NO	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.24%	3.69%	Pass
15	NO	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.18%	3.17%	Pass
15	NO	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.27% S	1.00%	Pass
15	NO	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.36% S	1.03%	Pass
15	NO	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.36% S	0.38%	Pass
16	PL	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.22%	1.64%	Pass
16	PL	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
16	PL	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.38%	2.33%	Pass
16	PL	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.96% S	1.05%	Pass
16	PL	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.50% S	3.49%	Pass
16	PL	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.33% S	0.52%	Pass
16	PL	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.25% S	1.57%	Pass
16	PL	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.71%	5.86%	Pass
16	PL	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.28%	2.02%	Pass
16	PL	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.78% S	0.54%	Pass
16	PL	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.36%	1.24%	Pass
16	PL	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.00%	2.73%	Pass
16	PL	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.60% S	0.75%	Pass
16	PL	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.03%	0.92%	Pass
16	PL	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.30% S	2.17%	Pass
16	PL	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	18.92% S	5.47%	
16	PL	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.91% S	1.11%	Pass
16	PL	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.94%	4.05%	Pass
16	PL	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.86% S	1.00%	Pass
16	PL	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.86% S	1.80%	Pass
16	PL	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.10% S	1.10%	Pass
18	RO	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.95%	19.54%	No pass
18	RO	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	135.59% S	39.45%	No pass
18	RO	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.32%	3.95%	Pass
18	RO	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.73% S	1.57%	Pass
18	RO	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
18	RO	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.49% S	0.20%	Pass
18	RO	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	326.58% S	73.91%	
18	RO	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
18	RO	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
18	RO	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.19% S	0.51%	Pass
19	ES	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	13.61% S	4.32%	Pass

19	ES	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.93%	7.31%	No pass
19	ES	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.58% S	1.57%	Pass
19	ES	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.75%	5.41%	Pass
19	ES	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.26% S	1.77%	Pass
19	ES	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.56%	1.55%	Pass
19	ES	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.77% S	2.79%	Pass
19	ES	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.03%	0.82%	Pass
19	ES	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.30%	0.89%	Pass
19	ES	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.58%	1.15%	Pass
19	ES	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	20.94% S	12.12%	
19	ES	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.39% S	0.74%	Pass
19	ES	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.19% S	2.82%	Pass
19	ES	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
19	ES	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.50%	1.67%	Pass
20	SE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.50%	1.19%	Pass
20	SE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.82% S	1.31%	Pass
20	SE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.59% S	1.17%	Pass
20	SE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.04% S	0.52%	Pass
20	SE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.45% S	1.28%	Pass
20	SE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.00% S	1.38%	Pass
20	SE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.64% S	1.21%	Pass
20	SE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.89% S	0.84%	Pass
20	SE	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.74%	9.18%	Pass
20	SE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.06% S	0.79%	Pass
20	SE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.12%	2.36%	Pass
20	SE	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.38%	2.15%	Pass
20	SE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.94% S	0.68%	Pass
20	SE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.17%	0.51%	Pass
20	SE	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.00%	1.74%	Pass
20	SE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.25%	4.39%	
20	SE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.23%	2.77%	Pass
20	SE	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.11%	1.76%	Pass

20	SE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.30%	5.45%	Pass
20	SE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.43%	1.03%	Pass
20	SE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.60%	1.58%	Pass
21	CH	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.91%	1.94%	Pass
21	CH	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.06%	2.09%	Pass
21	CH	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.89% S	1.15%	Pass
21	CH	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.18%	3.22%	Pass
21	CH	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.90%	2.48%	Pass
21	CH	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.97%	1.12%	Pass
21	CH	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	16.15% S	4.35%	
21	CH	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.20%	0.74%	Pass
21	CH	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.64% S	0.75%	Pass
21	CH	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.55%	1.54%	Pass
23	COM	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.86% S	3.19%	Pass
23	COM	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.27%	6.53%	Pass
23	COM	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.34% S	4.45%	Pass
23	COM	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.14%	1.94%	Pass
23	COM	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.54%	10.31%	No pass
23	COM	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.36% S	1.33%	Pass
23	COM	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.27%	0.63%	Pass
23	COM	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-48.66%	48.31%	
23	COM	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-12.69% S	6.65%	Pass
23	COM	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.23% S	3.51%	Pass
23	COM	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.92% S	2.19%	Pass
24	RS	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
24	RS	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
24	RS	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.17% S	2.02%	Pass
24	RS	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
24	RS	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
24	RS	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.13% S	3.37%	Pass
24	RS	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
24	RS	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
24	RS	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
24	RS	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.37%	3.47%	Pass
26	CA	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.51%	1.49%	Pass
26	CA	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.65%	0.78%	Pass
26	CA	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.37%	1.70%	Pass
26	CA	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
26	CA	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.14%	1.55%	Pass
26	CA	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00%	0.31%	Pass

26	CA	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.39%	4.16%	
26	CA	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.44% S	1.11%	Pass
26	CA	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.44% S	1.69%	Pass
26	CA	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.36%	0.77%	Pass
27	EDU	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.17%	1.49%	Pass
27	EDU	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.55%	2.09%	Pass
27	EDU	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.77%	1.97%	Pass
27	EDU	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.86% S	2.16%	Pass
27	EDU	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.75%	3.10%	Pass
27	EDU	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.67% S	0.41%	Pass
27	EDU	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	14.85% S	8.23%	
27	EDU	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.30%	2.03%	Pass
27	EDU	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.28% S	2.07%	Pass
27	EDU	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.01%	1.54%	Pass
30	EU	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.56%	6.56%	Pass
30	EU	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	11.76% S	0.00%	Pass
30	EU	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	11.89% S	9.86%	Pass
30	EU	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-11.56% S	4.17%	Pass
30	EU	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	9.38% S	4.97%	Pass
30	EU	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.71%	2.76%	Pass
30	EU	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	20.26% S	10.48%	
30	EU	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.76% S	1.29%	Pass
30	EU	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.49%	1.63%	Pass
30	EU	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.05% S	5.01%	No pass
31	SK	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.44% S	0.75%	Pass
31	SK	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.17% S	2.61%	Pass
31	SK	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.08%	0.77%	Pass
31	SK	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-12.09% S	0.52%	Pass
31	SK	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.51% S	2.94%	Pass
31	SK	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.04%	0.69%	Pass
31	SK	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.16%	3.51%	Pass
31	SK	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.35%	2.26%	Pass
31	SK	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.79%	3.25%	Pass
31	SK	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.18% S	2.18%	Pass
31	SK	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.40% S	1.86%	Pass
31	SK	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.61%	5.75%	Pass
31	SK	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.25%	0.82%	Pass
31	SK	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.30%	1.84%	Pass
31	SK	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.85% S	5.21%	Pass
31	SK	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.58%	15.06%	

31	SK	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-14.44% S	1.11%	Pass
31	SK	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.76%	5.15%	Pass
31	SK	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.86% S	1.56%	Pass
31	SK	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.95%	2.06%	Pass
31	SK	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.18%	2.41%	Pass
32	LT	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.34%	5.22%	Pass
32	LT	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-90.00% S	40.72%	No pass
32	LT	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	12.26% S	3.88%	Pass
32	LT	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-15.69% S	6.53%	Pass
32	LT	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.97%	2.02%	Pass
32	LT	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.12% S	2.92%	Pass
32	LT	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.02%	2.67%	Pass
32	LT	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	12.81% S	2.68%	Pass
32	LT	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.40%	6.29%	Pass
32	LT	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.44% S	0.85%	Pass
32	LT	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.97%	2.79%	Pass
32	LT	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.29%	2.12%	Pass
32	LT	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	97.89% S	3.67%	Pass
32	LT	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.44%	6.23%	No pass
32	LT	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.15% S	3.91%	Pass
32	LT	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.66% S	9.79%	
32	LT	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.66% S	0.74%	Pass
32	LT	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.18%	7.63%	Pass
32	LT	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.64%	1.69%	Pass
32	LT	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.16%	3.47%	Pass
32	LT	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-87.13% S	14.40%	Pass
33	LV	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.12% S	1.79%	Pass
33	LV	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-16.67%	32.66%	No pass
33	LV	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	10.51% S	7.78%	Pass
33	LV	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.07%	4.96%	Pass
33	LV	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.23% S	4.82%	Pass
33	LV	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.67% S	5.16%	Pass
33	LV	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.88% S	3.04%	Pass
33	LV	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.17%	12.56%	Pass
33	LV	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.10%	2.18%	Pass
33	LV	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	16.67% S	12.10%	Pass
33	LV	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.88%	7.45%	Pass
33	LV	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.69%	3.04%	Pass
33	LV	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
33	LV	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.12%	2.15%	Pass

33	LV	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.87%	4.13%	Pass
33	LV	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.59% S	7.27%	
33	LV	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.60% S	3.33%	Pass
33	LV	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.20%	2.28%	Pass
33	LV	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.35% S	2.13%	Pass
33	LV	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.54% S	1.93%	Pass
33	LV	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
35	HR	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.61% S	2.68%	Pass
35	HR	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.94% S	0.52%	Pass
35	HR	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.53% S	0.37%	Pass
35	HR	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.81% S	2.52%	Pass
35	HR	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.73%	1.55%	Pass
35	HR	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.85%	0.61%	Pass
35	HR	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	33.43% S	9.07%	
35	HR	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.42% S	0.92%	Pass
35	HR	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.86% S	0.63%	Pass
35	HR	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.31% S	1.03%	Pass
36	SI	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.75% S	0.89%	Pass
36	SI	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.50%	1.96%	Pass
36	SI	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.54%	5.45%	Pass
36	SI	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.14%	1.31%	Pass
36	SI	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.08% S	3.12%	Pass
36	SI	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-26.38% S	23.89%	Pass
36	SI	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.98%	1.61%	Pass
36	SI	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.28% S	0.67%	Pass
36	SI	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.28%	0.36%	Pass
36	SI	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.85%	1.24%	Pass
36	SI	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.64% S	0.41%	Pass
36	SI	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.27%	2.96%	Pass
36	SI	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00%	0.95%	
36	SI	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.85%	1.29%	Pass
36	SI	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.46%	1.88%	Pass
36	SI	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.36%	0.39%	Pass
36	SI	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.10% S	0.54%	Pass
38	EE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.53%	1.34%	Pass
38	EE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.17% S	3.70%	Pass
38	EE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-17.12% S	1.56%	Pass
38	EE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-13.63% S	3.66%	Pass
38	EE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.40%	5.46%	Pass
38	EE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			

38	EE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.41%	2.67%	Pass
38	EE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-56.44% S	4.19%	Pass
38	EE	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.44%	1.75%	Pass
38	EE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-19.11% S	3.93%	Pass
38	EE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.15%	3.73%	Pass
38	EE	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	17.58% S	6.92%	Pass
38	EE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-31.53% S	7.08%	Pass
38	EE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.76% S	0.72%	Pass
38	EE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-45.59% S	18.60%	
38	EE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.70%	1.66%	Pass
38	EE	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.60%	1.73%	Pass
38	EE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.83% S	1.88%	Pass
38	EE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.24%	0.51%	Pass
38	EE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.89% S	1.52%	Pass
39	PL	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.25%	1.64%	Pass
39	PL	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00%	0.00%	Pass
39	PL	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	14.30% S	3.88%	Pass
39	PL	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.17% S	0.78%	Pass
39	PL	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.02% S	3.40%	Pass
39	PL	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00% S	8.59%	Pass
39	PL	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.97%	1.87%	Pass
39	PL	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00% S	0.00%	Pass
39	PL	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-16.74%	14.17%	No pass
39	PL	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	10.10%	12.10%	Pass
39	PL	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.03% S	2.17%	Pass
39	PL	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.51%	4.01%	Pass
39	PL	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.25% S	6.80%	Pass
39	PL	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.51%	0.72%	Pass
39	PL	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	12.23% S	4.78%	
39	PL	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.38%	1.29%	Pass
39	PL	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.21%	6.66%	Pass
39	PL	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.89% S	0.50%	Pass
39	PL	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.76% S	1.67%	Pass
39	PL	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00% S	0.00%	Pass
41	NET	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	19.25% S	3.05%	Pass
41	NET	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
41	NET	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
41	NET	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.78% S	1.47%	Pass
41	NET	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.18%	15.97%	Pass
41	NET	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.50% S	15.38%	Pass

41	NET	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.94%	37.18%	No pass
42	COM	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
42	COM	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
42	COM	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
42	COM	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	15.93%	12.84%	No pass
42	COM	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
42	COM	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
42	COM	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.18%	6.08%	Pass
42	COM	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.54%	25.06%	
42	COM	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
42	COM	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
42	COM	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
45	COM	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.58%	13.42%	No pass
45	COM	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.82% S	1.57%	Pass
45	COM	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.47% S	1.42%	Pass
45	COM	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.14%	1.21%	Pass
45	COM	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.61% S	3.73%	Pass
45	COM	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.80% S	1.53%	Pass
45	COM	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.69%	9.89%	
45	COM	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.37% S	3.33%	Pass
45	COM	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.20%	8.57%	Pass
45	COM	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.97% S	0.51%	Pass
46	PL	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
46	PL	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
46	PL	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
46	PL	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.95% S	2.49%	Pass
46	PL	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
46	PL	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
46	PL	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.40% S	2.44%	
46	PL	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
46	PL	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
46	PL	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
48	BE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.00%	5.88%	Pass
48	BE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.03%	3.10%	Pass
48	BE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.63%	6.70%	Pass
48	BE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.49% S	4.98%	Pass
48	BE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.33%	1.33%	Pass
48	BE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.19% S	5.85%	Pass
48	BE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.64% S	2.59%	Pass
49	CY	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.27% S	3.88%	Pass

49	CY	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.57%	1.05%	Pass
49	CY	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.76% S	1.65%	Pass
49	CY	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.88%	2.82%	Pass
49	CY	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.33% S	4.35%	Pass
49	CY	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.38% S	0.92%	Pass
49	CY	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-51.58% S	1.90%	
49	CY	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.74%	2.77%	Pass
49	CY	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.37% S	1.00%	Pass
49	CY	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.31%	1.16%	Pass
50	FR	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.17% S	2.39%	Pass
50	FR	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-14.38% S	1.83%	Pass
50	FR	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.57% S	1.88%	Pass
50	FR	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.55%	2.40%	Pass
50	FR	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.00%	4.66%	Pass
50	FR	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.44% S	0.31%	Pass
50	FR	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	18.86% S	6.30%	
50	FR	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.04% S	1.85%	Pass
50	FR	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.41%	0.63%	Pass
50	FR	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.37% S	0.64%	Pass
51	COM	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	40.00%	1.74%	Pass
51	COM	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
51	COM	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
51	COM	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.74% S	1.55%	Pass
51	COM	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.40%	1.27%	Pass
51	COM	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.29%	0.82%	Pass
51	COM	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.46%	2.71%	Pass
52	PL	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.92% S	0.87%	Pass
52	PL	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-11.23% S	2.34%	Pass
52	PL	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
52	PL	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
52	PL	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.36% S	2.42%	Pass
52	PL	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-12.25%	7.48%	Pass
52	PL	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
53	IT	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.33% S	4.35%	Pass
53	IT	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.59% S	3.91%	Pass
53	IT	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.25%	1.72%	Pass
53	IT	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.52% S	2.93%	Pass
53	IT	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.75% S	13.91%	Pass
53	IT	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-16.14% S	2.04%	Pass
53	IT	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	17.95% S	5.44%	Pass

110	DE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.01%	3.28%	Pass
110	DE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.33% S	6.53%	Pass
110	DE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	15.12% S	15.55%	Pass
110	DE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.03% S	2.35%	Pass
110	DE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.87%	4.50%	Pass
110	DE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.33% S	8.59%	Pass
110	DE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.17%	6.18%	Pass
110	DE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	99.75% S	98.40%	No pass
110	DE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	78.61% S	75.60%	No pass
110	DE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.63% S	0.93%	Pass
110	DE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.75% S	4.76%	Pass
110	DE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-93.03%	64.99%	No pass
110	DE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.78%	12.20%	
110	DE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.37% S	1.29%	Pass
110	DE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.81% S	4.19%	Pass
110	DE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.12% S	5.66%	No pass
110	DE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.88%	4.35%	Pass
114	IT	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.99%	3.13%	Pass
114	IT	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
114	IT	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
114	IT	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	24.18% S	5.49%	Pass
114	IT	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.71%	4.04%	Pass
114	IT	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	12.50% S	17.19%	Pass
114	IT	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.26%	2.16%	Pass
114	IT	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	15.66% S	12.56%	Pass
114	IT	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-39.58%	24.19%	Pass
114	IT	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.50% S	9.62%	Pass
114	IT	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00% S	0.00%	Pass
114	IT	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.27%	0.61%	Pass
114	IT	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-18.63% S	12.97%	
114	IT	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.56%	2.96%	Pass
114	IT	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.59%	1.88%	Pass
114	IT	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.78%	2.96%	Pass
114	IT	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.83%	1.63%	Pass
115	DE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.94% S	1.79%	Pass
115	DE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	18.00% S	2.83%	Pass
115	DE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.87% S	3.10%	Pass
115	DE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.50% S	2.87%	Pass
115	DE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.45% S	3.44%	Pass
115	DE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.96% S	0.69%	Pass

115	DE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.62% S	1.98%	Pass
115	DE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.78% S	0.21%	Pass
115	DE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.99% S	1.33%	Pass
115	DE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.61% S	1.55%	Pass
115	DE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.56% S	1.09%	Pass
115	DE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.21% S	1.53%	Pass
115	DE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	36.47% S	9.46%	
115	DE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.64% S	2.77%	Pass
115	DE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.81%	1.44%	Pass
115	DE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.91% S	1.41%	Pass
115	DE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.33% S	1.31%	Pass
116	CH	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.06%	1.79%	Pass
116	CH	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.68%	7.32%	Pass
116	CH	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.80% S	21.77%	No pass
116	CH	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.82%	4.76%	Pass
116	CH	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.88%	4.04%	Pass
116	CH	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.56% S	26.08%	No pass
116	CH	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	39.65%	34.39%	
116	CH	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.43% S	2.96%	Pass
116	CH	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.43% S	0.19%	Pass
116	CH	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.47% S	26.17%	No pass
117	DE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.25% S	0.30%	Pass
117	DE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
117	DE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.53%	4.67%	Pass
117	DE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.62% S	1.31%	Pass
117	DE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-19.85% S	9.96%	Pass
117	DE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
117	DE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.23%	14.78%	No pass
117	DE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-14.39% S	2.51%	Pass
117	DE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-16.59% S	1.21%	Pass
117	DE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.41% S	3.42%	Pass
117	DE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
117	DE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.10% S	0.61%	Pass
117	DE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	13.69% S	18.93%	
117	DE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-16.75% S	3.33%	Pass
117	DE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.81% S	4.19%	Pass
117	DE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-45.94% S	25.20%	No pass
117	DE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	27.90% S	6.80%	Pass
118	DE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.12% S	3.28%	Pass
118	DE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			

118	DE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-40.54% S	1.58%	Pass
118	DE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-37.91% S	9.67%	Pass
118	DE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.24% S	2.89%	Pass
118	DE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
118	DE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.84%	3.07%	Pass
118	DE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
118	DE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.25% S	3.63%	Pass
118	DE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-27.45% S	5.59%	Pass
118	DE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
118	DE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.64%	1.64%	Pass
118	DE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.28%	6.44%	
118	DE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.16% S	2.59%	Pass
118	DE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-11.52% S	3.88%	Pass
118	DE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.03%	2.31%	Pass
118	DE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.56%	5.44%	Pass
120	DE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	12.99% S	2.09%	Pass
120	DE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.17%	4.79%	Pass
120	DE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.43%	5.45%	Pass
120	DE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.28% S	0.26%	Pass
120	DE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.97% S	3.76%	Pass
120	DE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	16.67% S	5.50%	Pass
120	DE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.11% S	2.05%	Pass
120	DE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.61% S	0.67%	Pass
120	DE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.61%	1.57%	Pass
120	DE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.58%	4.66%	Pass
120	DE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.20% S	1.50%	Pass
120	DE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.24% S	2.25%	Pass
120	DE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.15% S	1.98%	
120	DE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-14.10% S	3.33%	Pass
120	DE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.26% S	0.94%	Pass
120	DE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.22% S	1.80%	Pass
120	DE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.97% S	1.09%	Pass
121	DE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.96%	2.24%	Pass
121	DE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-26.83% S	19.38%	Pass
121	DE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.26% S	0.78%	Pass
121	DE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-15.28% S	4.44%	Pass
121	DE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-11.40% S	7.89%	Pass
121	DE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.54% S	0.34%	Pass
121	DE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.36%	4.54%	Pass
121	DE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-14.97% S	2.55%	Pass

121	DE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.60%	0.79%	Pass
121	DE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.96% S	1.24%	Pass
121	DE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-17.82% S	2.11%	Pass
121	DE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.02% S	0.41%	Pass
121	DE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.59%	5.70%	
121	DE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.39% S	5.54%	Pass
121	DE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.34% S	4.32%	Pass
121	DE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.46% S	2.19%	Pass
121	DE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.08% S	1.31%	Pass
124	BE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
124	BE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.05%	7.05%	Pass
124	BE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.45% S	1.19%	Pass
124	BE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
124	BE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.03%	3.42%	Pass
124	BE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.15%	1.53%	Pass
124	BE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-14.80% S	3.61%	
124	BE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.46%	14.78%	Pass
124	BE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.61%	0.63%	Pass
124	BE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.90%	2.19%	Pass
125	DE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.87% S	2.39%	Pass
125	DE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.25%	3.29%	Pass
125	DE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.88%	0.79%	Pass
125	DE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.53% S	1.57%	Pass
125	DE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.99% S	1.01%	Pass
125	DE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.46%	1.55%	Pass
125	DE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.17% S	7.17%	No pass
125	DE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.24%	4.61%	Pass
125	DE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.39% S	0.79%	Pass
125	DE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.78% S	2.33%	Pass
125	DE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.47% S	1.22%	Pass
125	DE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.83% S	0.72%	Pass
125	DE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.33%	19.79%	
125	DE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.81%	1.29%	Pass
125	DE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.79% S	3.07%	Pass
125	DE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.42% S	2.96%	Pass
125	DE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	15.83% S	1.31%	Pass
126	IT	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-18.79% S	7.61%	No pass
126	IT	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	87.27% S	15.68%	No pass
126	IT	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.89% S	2.34%	Pass
126	IT	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.49%	10.03%	No pass

126	IT	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	44.10% S	10.56%	No pass
126	IT	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.58%	3.37%	Pass
126	IT	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	84.09% S	7.52%	
126	IT	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.69%	13.12%	Pass
126	IT	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.06%	1.82%	Pass
126	IT	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.80% S	1.03%	Pass
141	JP	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
141	JP	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-18.92% S	2.32%	Pass
141	JP	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
141	JP	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
141	JP	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
141	JP	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
141	JP	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
145	EE	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.99%	10.19%	No pass
145	EE	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.17%	5.01%	Pass
145	EE	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.94% S	2.36%	Pass
145	EE	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	23.28% S	6.01%	Pass
145	EE	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.94% S	3.12%	Pass
145	EE	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
145	EE	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.42% S	1.43%	Pass
145	EE	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.16%	0.42%	Pass
145	EE	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.47% S	0.91%	Pass
145	EE	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.44%	4.13%	Pass
145	EE	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.39% S	1.36%	Pass
145	EE	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.17%	1.64%	Pass
145	EE	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-11.84% S	8.39%	
145	EE	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-13.46% S	1.29%	Pass
145	EE	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.04% S	1.00%	Pass
145	EE	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.11% S	1.41%	Pass
145	EE	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.34% S	1.14%	Pass
146	LU	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.33% S	0.00%	Pass
146	LU	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-11.67% S	3.13%	Pass
146	LU	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.50% S	7.39%	Pass
146	LU	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.99%	2.14%	Pass
146	LU	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.86%	0.97%	Pass
146	LU	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.96% S	0.95%	Pass
146	LU	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.82% S	28.25%	No pass
153	SI	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	19.11% S	4.32%	Pass
153	SI	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.89%	8.36%	Pass
153	SI	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.01% S	3.99%	Pass

153	SI	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.88% S	4.06%	Pass
153	SI	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	9.46% S	8.07%	Pass
153	SI	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.17% S	2.04%	Pass
153	SI	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	32.14%	237.04%	
153	SI	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.31%	2.59%	Pass
153	SI	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.28%	3.51%	Pass
153	SI	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.54% S	1.03%	Pass
155	UK	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
155	UK	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.85% S	1.05%	Pass
155	UK	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.50% S	0.18%	Pass
155	UK	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.09%	1.72%	Pass
155	UK	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.24%	0.93%	Pass
155	UK	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.49%	1.02%	Pass
155	UK	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.59%	5.91%	
155	UK	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.27% S	1.85%	Pass
155	UK	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.04% S	1.63%	Pass
155	UK	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.99% S	1.80%	Pass
158	ASIA	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.30% S	2.09%	Pass
158	ASIA	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.60% S	1.83%	Pass
158	ASIA	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.34%	1.01%	Pass
158	ASIA	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.83%	2.52%	Pass
158	ASIA	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.97%	21.06%	No pass
158	ASIA	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.09% S	1.55%	Pass
158	ASIA	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.38%	5.17%	Pass
158	ASIA	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.02% S	2.25%	Pass
158	ASIA	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-14.80%	13.61%	
158	ASIA	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.29% S	0.74%	Pass
158	ASIA	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.16%	23.69%	No pass
158	ASIA	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.19%	0.75%	Pass
158	ASIA	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.71% S	0.39%	Pass
166	PL	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.75% S	2.09%	Pass
166	PL	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
166	PL	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	21.45%	19.42%	Pass
166	PL	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.33%	4.96%	Pass
166	PL	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.24% S	0.96%	Pass
166	PL	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
166	PL	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.86% S	2.74%	Pass
166	PL	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.80%	16.20%	Pass
166	PL	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-31.99% S	17.60%	Pass
166	PL	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.61%	3.42%	Pass

166	PL	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
166	PL	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.31%	0.72%	Pass
166	PL	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
166	PL	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.64%	2.96%	Pass
166	PL	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.03%	8.83%	Pass
166	PL	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.12%	0.64%	Pass
166	PL	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.81% S	1.61%	Pass
169	UK	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	41.33% S	4.57%	Pass
169	UK	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.87%	2.31%	Pass
169	UK	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	23.00%	10.66%	Pass
169	UK	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.43%	0.67%	Pass
169	UK	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.17%	0.79%	Pass
169	UK	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.63%	2.04%	Pass
169	UK	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	9.20% S	1.59%	Pass
171	FR	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.25% S	3.27%	Pass
171	FR	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	5.91% S	1.39%	Pass
171	FR	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.58% S	2.27%	Pass
171	FR	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.38% S	1.37%	Pass
171	FR	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.28% S	0.14%	Pass
171	FR	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.20% S	0.27%	Pass
171	FR	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.92% S	1.91%	Pass
174	COM	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.12% S	3.88%	Pass
174	COM	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
174	COM	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-48.21% S	14.79%	Pass
174	COM	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	35.62% S	3.40%	Pass
174	COM	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-15.08% S	4.40%	Pass
174	COM	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
174	COM	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.61%	2.20%	Pass
174	COM	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-42.93% S	18.84%	Pass
174	COM	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-91.77%	56.14%	No pass
174	COM	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-30.63% S	14.52%	Pass
174	COM	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	22.10%	16.14%	No pass
174	COM	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
174	COM	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.74% S	4.29%	Pass
174	COM	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.59%	16.50%	No pass
174	COM	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-31.60% S	5.13%	
174	COM	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	19.53% S	10.34%	Pass
174	COM	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-93.65%	56.25%	No pass
174	COM	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.49% S	1.19%	Pass
174	COM	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.07% S	0.77%	Pass

174	COM	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
178	RU	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	27.70% S	17.60%	No pass
178	RU	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.83%	4.35%	Pass
178	RU	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.54%	15.56%	Pass
178	RU	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-18.14%	16.46%	No pass
178	RU	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-15.06% S	4.63%	Pass
178	RU	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.42%	19.94%	Pass
178	RU	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.64%	2.78%	Pass
178	RU	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.02%	8.37%	Pass
178	RU	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.66%	2.04%	Pass
178	RU	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.80%	5.44%	Pass
178	RU	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.21%	2.79%	Pass
178	RU	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.14%	5.44%	Pass
178	RU	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	323.93% S	169.12%	No pass
178	RU	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	28.82% S	9.57%	
178	RU	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.95% S	1.66%	Pass
178	RU	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	691.05% S	654.81%	No pass
178	RU	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.74% S	1.44%	Pass
178	RU	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	198.80% S	56.19%	No pass
178	RU	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.05%	7.62%	Pass
179	RU	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.65%	5.07%	Pass
179	RU	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
179	RU	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	46.43%	0.00%	Pass
179	RU	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	22.88% S	7.58%	Pass
179	RU	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.51% S	3.85%	Pass
179	RU	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
179	RU	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.64%	2.31%	Pass
179	RU	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.33%	0.42%	Pass
179	RU	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
179	RU	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-28.79% S	7.76%	Pass
179	RU	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
179	RU	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.02%	1.12%	Pass
179	RU	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
179	RU	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.06%	4.62%	Pass
179	RU	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.80% S	3.51%	Pass
179	RU	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.82%	4.50%	Pass
179	RU	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-13.88% S	0.54%	Pass
181	NL	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	14.68% S	1.49%	Pass
181	NL	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.33%	16.55%	Pass
181	NL	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-9.97%	16.33%	Pass

181	NL	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	15.01% S	131.68%	No pass
181	NL	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.17%	6.10%	Pass
181	NL	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-50.00% S	0.00%	Pass
181	NL	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	15.99%	85.25%	No pass
181	NL	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.63% S	33.08%	No pass
181	NL	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.74%	1.21%	Pass
181	NL	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.46%	0.00%	Pass
181	NL	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.50%	5.44%	Pass
181	NL	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.86%	3.37%	Pass
181	NL	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-4.50%	3.71%	
181	NL	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.91%	4.62%	Pass
181	NL	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.64%	2.63%	Pass
181	NL	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.14%	1.29%	Pass
181	NL	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.92%	2.72%	Pass
183	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.67%	3.91%	Pass
184	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.50%	2.61%	Pass
185	ORG	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
185	ORG	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
185	ORG	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
185	ORG	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
185	ORG	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
185	ORG	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
185	ORG	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.87%	2.17%	Pass
185	ORG	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.18%	2.17%	Pass
186	RS	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
186	RS	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-42.42% S	9.95%	Pass
186	RS	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
187	COM	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.54%	1.09%	Pass
188	COM	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.19%	1.74%	Pass
189	RS	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
189	RS	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
189	RS	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			

189	RS	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
189	RS	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
189	RS	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
189	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00%	1.74%	Pass
189	RS	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
192	COM	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.19% S	1.30%	Pass
193	RS	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.67%	3.28%	Pass
193	RS	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00% S	0.00%	Pass
193	RS	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	12.74% S	15.56%	Pass
193	RS	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00%	0.00%	Pass
193	RS	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-8.71% S	2.09%	Pass
193	RS	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.18%	0.65%	Pass
193	RS	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
193	RS	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	25.38% S	10.42%	No pass
193	RS	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.62%	1.09%	Pass
194	RS	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
194	RS	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.08%	8.88%	Pass
194	RS	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
194	RS	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
194	RS	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.17% S	0.62%	Pass
194	RS	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
194	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.40%	0.56%	Pass
194	RS	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
194	RS	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-7.11% S	1.48%	Pass
194	RS	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-5.02% S	3.63%	Pass
194	RS	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
196	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.18%	1.30%	Pass
197	COM	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.63%	3.43%	Pass
197	COM	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
197	COM	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			
197	COM	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf			

197	COM	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
197	COM	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
197	COM	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.01% S	0.65%		Pass
197	COM	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
197	COM	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
197	COM	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
197	COM	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
198	RS	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
198	RS	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
198	RS	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
198	RS	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
198	RS	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
198	RS	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
198	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	1.09%	2.39%		Pass
198	RS	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
200	CH	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.45%	3.73%		Pass
200	CH	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.76% S	2.09%		Pass
200	CH	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-1.62% S	1.24%		Pass
200	CH	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.64% S	3.51%		Pass
200	CH	HNO3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.62%	2.59%		Pass
200	CH	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	4.23% S	2.48%		Pass
200	CH	NH3-N on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-6.70%	5.01%		Pass
200	CH	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.73%	1.02%		Pass
200	CH	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.00% S	1.30%		Pass
200	CH	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.87%	16.07%		
200	CH	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	7.04% S	2.59%		Pass
200	CH	SO2-S on impregnated filter	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-3.61%	3.61%		Pass
200	CH	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-2.06% S	1.38%		Pass
200	CH	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.48%	1.03%		Pass
201	RS	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.37% S	1.94%		Pass
201	RS	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-36.93% S	12.54%		Pass
201	RS	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	237.07% S	166.89%		No pass
201	RS	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-19.21% S	4.04%		Pass
201	RS	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				

201	RS	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-0.24%	2.61%		Pass
201	RS	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-68.34% S	6.32%		
201	RS	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
201	RS	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-24.86% S	1.63%		Pass
203	RS	Ammonium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Calcium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Chloride in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Conductivity in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Magnesium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-20.27% S	7.14%		Pass
203	RS	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Nitrate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	3.02% S	1.30%		Pass
203	RS	pH in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Potassium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-10.00% S	6.10%		Pass
203	RS	Sodium in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	-14.63% S	11.58%		No pass
203	RS	Sulphate in precipitation	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
203	RS	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf				
205	DK	Arsenic	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.25%	2.61%		Pass
205	DK	Cadmium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	6.75% S	2.63%		Pass
205	DK	Chromium	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	0.25%	0.88%		Pass
205	DK	Copper	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.61% S	0.67%		Pass
205	DK	Lead	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	19.04% S	0.91%		Pass
205	DK	Nickel	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	2.73% S	1.36%		Pass
205	DK	Zinc	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	10.24% S	1.58%		Pass
206	COM	NO2-N in absorbing solution	EMEP37	20191206	http://www.nilu.no/projects/ccc/qameasure/emep37.pdf	8.69% S	6.51%		Pass