

Lab number	Lab name	Component	QA measure ID	QA date	QA document url	QA bias	QA variability	QA outcome
3	CZ	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.95% S	0.85%	Pass
3	CZ	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.71% S	7.85%	Pass
3	CZ	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.31%	3.91%	Pass
3	CZ	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.71% S	1.08%	Pass
3	CZ	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.74% S	0.33%	Pass
3	CZ	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.95% S	1.24%	Pass
3	CZ	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.94%	42.14%	No pass
3	CZ	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.47%	1.78%	Pass
3	CZ	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.28%	2.95%	Pass
3	CZ	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.29%	5.84%	Pass
3	CZ	NH3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.11%	1.72%	Pass
3	CZ	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.73%	1.97%	Pass
3	CZ	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.50% S	0.95%	Pass
3	CZ	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.47% S	0.39%	Pass
3	CZ	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.70% S	0.86%	Pass
3	CZ	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	26.72% S	28.66%	
3	CZ	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.13%	1.19%	Pass
3	CZ	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.05% S	0.68%	Pass
4	DK	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.59%	1.69%	Pass
4	DK	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.12% S	1.21%	Pass
4	DK	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.41% S	0.00%	Pass
4	DK	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.08% S	0.62%	Pass
4	DK	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
4	DK	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.00%	1.13%	Pass
4	DK	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.25% S	14.99%	Pass
4	DK	HNO3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.98% S	0.84%	Pass
4	DK	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.15% S	3.35%	Pass
4	DK	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.20% S	1.45%	Pass
4	DK	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.94% S	0.68%	Pass
4	DK	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.41% S	1.69%	Pass
4	DK	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.01%	3.36%	Pass
4	DK	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.41% S	0.52%	Pass
4	DK	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.03%	10.16%	Pass
4	DK	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.14%	0.98%	
4	DK	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.06% S	0.73%	Pass
4	DK	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.06% S	5.20%	No pass
4	DK	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.87%	8.59%	Pass
5	FI	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.23% S	2.31%	Pass
5	FI	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.73% S	0.60%	Pass
5	FI	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.22%	1.78%	Pass

5	FI	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.74% S	0.34%	Pass
5	FI	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.29% S	0.65%	Pass
5	FI	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.67% S	0.93%	Pass
5	FI	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.57% S	1.97%	Pass
5	FI	HNO3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.87% S	2.37%	Pass
5	FI	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.12% S	1.56%	Pass
5	FI	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.02% S	0.34%	Pass
5	FI	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.58% S	0.56%	Pass
5	FI	NH3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.94% S	3.77%	Pass
5	FI	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.53% S	0.56%	Pass
5	FI	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.96% S	1.60%	Pass
5	FI	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.79% S	0.79%	Pass
5	FI	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.78% S	1.48%	Pass
5	FI	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.18% S	5.27%	
5	FI	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.23%	1.50%	Pass
5	FI	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.48% S	0.59%	Pass
5	FI	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.77% S	0.46%	Pass
6	FR	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	47.75% S	9.43%	Pass
6	FR	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.57% S	0.27%	Pass
6	FR	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
6	FR	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.11% S	0.37%	Pass
6	FR	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.33%	2.20%	Pass
6	FR	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	13.18%	0.00%	Pass
6	FR	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.72% S	0.00%	Pass
6	FR	HNO3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.06%	1.89%	Pass
6	FR	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.57% S	3.01%	Pass
6	FR	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.93% S	0.41%	Pass
6	FR	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.89% S	1.04%	Pass
6	FR	NH3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.29% S	1.32%	Pass
6	FR	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.04%	2.83%	Pass
6	FR	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
6	FR	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.84% S	0.46%	Pass
6	FR	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.13% S	0.53%	Pass
6	FR	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.48%	27.56%	
6	FR	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.18% S	1.57%	Pass
6	FR	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.65%	0.37%	Pass
6	FR	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
8	DE	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.15% S	2.42%	Pass
8	DE	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.20% S	2.41%	Pass
8	DE	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.48% S	1.39%	Pass
8	DE	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.37% S	0.25%	Pass

8	DE	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.92% S	1.04%	Pass
8	DE	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.69% S	0.57%	Pass
8	DE	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.38% S	2.25%	Pass
8	DE	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.90% S	0.57%	Pass
8	DE	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.98% S	0.45%	Pass
8	DE	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.49% S	1.83%	Pass
8	DE	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.60% S	0.77%	Pass
8	DE	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	15.24% S	2.86%	Pass
8	DE	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.49% S	0.53%	Pass
8	DE	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.56%	1.79%	Pass
8	DE	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.21%	2.88%	Pass
8	DE	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.54% S	1.39%	Pass
8	DE	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-14.80% S	4.62%	
8	DE	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.05%	0.79%	Pass
8	DE	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.96% S	0.74%	Pass
8	DE	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.42%	1.32%	Pass
10	HU	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.15%	22.60%	Pass
10	HU	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-19.39% S	6.04%	Pass
10	HU	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-57.77% S	0.00%	Pass
10	HU	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.01% S	1.27%	Pass
10	HU	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.46% S	0.44%	Pass
10	HU	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
10	HU	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
10	HU	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.77% S	1.74%	Pass
10	HU	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.15% S	1.34%	Pass
10	HU	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.15% S	1.08%	Pass
10	HU	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.47%	1.57%	Pass
10	HU	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.53% S	4.78%	Pass
10	HU	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.37% S	3.89%	Pass
10	HU	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
10	HU	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.85%	8.96%	Pass
10	HU	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.54% S	2.62%	Pass
10	HU	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-68.29% S	38.80%	No pass
10	HU	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	13.51%	18.30%	
10	HU	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.76% S	1.42%	Pass
10	HU	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.38% S	1.34%	Pass
10	HU	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
12	IE	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-19.39% S	6.64%	Pass
12	IE	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.72% S	0.31%	Pass
12	IE	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.67% S	0.22%	Pass
12	IE	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.62% S	2.01%	Pass

12	IE	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-15.40% S	3.39%	Pass
12	IE	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.26% S	1.05%	Pass
12	IE	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-33.53% S	9.42%	No pass
12	IE	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.25% S	1.15%	Pass
12	IE	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.46% S	2.88%	Pass
12	IE	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-43.08% S	7.50%	
12	IE	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-18.82% S	0.89%	Pass
15	NO	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.11% S	3.14%	Pass
15	NO	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.25% S	0.91%	Pass
15	NO	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.63%	3.66%	Pass
15	NO	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.60% S	0.53%	Pass
15	NO	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.87%	2.40%	Pass
15	NO	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.90%	1.81%	Pass
15	NO	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.45%	1.75%	Pass
15	NO	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.01%	1.34%	Pass
15	NO	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.04%	2.23%	Pass
15	NO	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.48% S	0.34%	Pass
15	NO	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.88% S	1.61%	Pass
15	NO	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.18% S	2.12%	Pass
15	NO	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	18.43% S	3.58%	Pass
15	NO	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.23%	1.89%	Pass
15	NO	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.09% S	0.92%	Pass
15	NO	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.16% S	0.52%	Pass
15	NO	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.63%	1.45%	Pass
15	NO	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.16% S	12.54%	
15	NO	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.44%	0.63%	Pass
15	NO	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.15% S	0.89%	Pass
15	NO	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.58% S	1.24%	Pass
16	PL	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
16	PL	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.84%	1.81%	Pass
16	PL	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	42.99% S	20.55%	Pass
16	PL	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.72% S	0.43%	Pass
16	PL	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.75% S	1.71%	Pass
16	PL	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-23.07% S	8.32%	Pass
16	PL	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-36.64% S	10.24%	Pass
16	PL	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.39% S	2.16%	Pass
16	PL	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.67% S	1.56%	Pass
16	PL	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.82%	1.83%	Pass
16	PL	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.96% S	1.09%	Pass
16	PL	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.29%	3.89%	Pass
16	PL	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.06% S	2.53%	Pass

16	PL	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-34.16% S	16.18%	Pass
16	PL	NO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.79% S	0.92%	Pass
16	PL	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.65% S	0.92%	Pass
16	PL	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.18%	11.10%	Pass
16	PL	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.16%	4.74%	
16	PL	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.15%	2.06%	Pass
16	PL	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.44%	1.19%	Pass
16	PL	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-39.24% S	26.11%	No pass
19	ES	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.26%	2.72%	Pass
19	ES	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.32% S	1.86%	Pass
19	ES	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.92% S	0.87%	Pass
19	ES	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.10% S	1.56%	Pass
19	ES	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.00% S	0.34%	Pass
19	ES	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.36% S	0.60%	Pass
19	ES	NH3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.83% S	3.26%	Pass
19	ES	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.17%	1.67%	Pass
19	ES	NO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.23% S	0.69%	Pass
19	ES	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.10% S	1.57%	Pass
19	ES	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	58.49% S	6.20%	
19	ES	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.02%	2.08%	Pass
20	SE	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.14% S	1.09%	Pass
20	SE	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.26% S	1.21%	Pass
20	SE	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.03%	1.77%	Pass
20	SE	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.39%	0.62%	Pass
20	SE	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.82% S	1.31%	Pass
20	SE	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.08% S	0.57%	Pass
20	SE	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.07% S	0.50%	Pass
20	SE	HNO3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.90% S	1.72%	Pass
20	SE	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.92% S	0.67%	Pass
20	SE	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.56% S	1.16%	Pass
20	SE	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.70% S	2.18%	Pass
20	SE	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.11%	0.83%	Pass
20	SE	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.70%	1.58%	Pass
20	SE	NO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.89% S	0.92%	Pass
20	SE	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.20% S	0.92%	Pass
20	SE	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.95% S	5.34%	Pass
20	SE	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.16%	12.10%	
20	SE	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.66% S	0.95%	Pass
20	SE	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.51% S	0.89%	Pass
20	SE	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	12.02% S	1.36%	Pass
21	CH	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-21.92% S	1.81%	Pass

21	CH	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.11% S	0.28%	Pass
21	CH	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.19% S	0.47%	Pass
21	CH	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.80% S	1.12%	Pass
21	CH	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.26% S	0.37%	Pass
21	CH	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.74% S	0.77%	Pass
21	CH	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.40%	2.25%	Pass
21	CH	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.42% S	0.79%	Pass
21	CH	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.71% S	2.71%	
21	CH	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.80% S	0.59%	Pass
22	RU	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
22	RU	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.11%	14.79%	Pass
22	RU	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	142.12% S	23.58%	Pass
22	RU	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.68% S	1.89%	Pass
22	RU	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	22.12% S	7.05%	No pass
22	RU	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
22	RU	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	75.34% S	15.41%	Pass
22	RU	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	20.50% S	4.45%	Pass
22	RU	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.43%	7.14%	Pass
22	RU	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.66% S	5.93%	Pass
22	RU	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.18%	5.64%	Pass
22	RU	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.47% S	8.06%	No pass
22	RU	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
22	RU	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.65% S	2.99%	Pass
22	RU	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.31% S	1.57%	Pass
22	RU	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	15.03%	15.51%	Pass
22	RU	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.04%	7.00%	
22	RU	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-16.94% S	3.13%	Pass
22	RU	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.16% S	0.59%	Pass
22	RU	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
23	GB	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-11.37% S	9.66%	Pass
23	GB	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.70% S	1.30%	Pass
23	GB	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.00% S	0.33%	Pass
23	GB	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.63%	5.35%	Pass
23	GB	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-18.23% S	2.76%	Pass
23	GB	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.45% S	1.17%	Pass
23	GB	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.23% S	3.89%	Pass
23	GB	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.30% S	2.36%	Pass
23	GB	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-14.80% S	4.90%	
23	GB	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-14.92% S	1.34%	Pass
26	CA	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.00%	1.21%	Pass
26	CA	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.14% S	0.22%	Pass

26	CA	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf				
26	CA	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.39% S	1.56%	Pass	
26	CA	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.46%	0.78%	Pass	
26	CA	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.01% S	0.64%	Pass	
26	CA	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.00%	1.36%	Pass	
26	CA	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.98% S	1.05%	Pass	
26	CA	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-33.10% S	21.22%		
26	CA	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.45%	0.59%	Pass	
27	US	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.48%	3.62%	Pass	
27	US	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.69% S	0.97%	Pass	
27	US	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.83% S	0.94%	Pass	
27	US	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.80% S	0.89%	Pass	
27	US	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.62%	2.54%	Pass	
27	US	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.24% S	1.48%	Pass	
27	US	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.14% S	1.11%	Pass	
27	US	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.27%	0.92%	Pass	
27	US	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.88% S	3.90%		
27	US	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.31%	1.34%	Pass	
30	IT	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.22% S	3.92%	Pass	
30	IT	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.81% S	0.87%	Pass	
30	IT	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.04% S	1.02%	Pass	
30	IT	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.39% S	4.68%	Pass	
30	IT	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.18% S	1.45%	Pass	
30	IT	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.63% S	0.48%	Pass	
30	IT	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.15%	2.47%	Pass	
30	IT	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.85%	1.70%	Pass	
30	IT	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	10.70%	22.11%		
30	IT	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.61% S	0.45%	Pass	
31	SK	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	71.60%	42.43%	No pass	
31	SK	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.55% S	0.91%	Pass	
31	SK	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.41%	54.22%	No pass	
31	SK	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.06% S	0.84%	Pass	
31	SK	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.17% S	1.64%	Pass	
31	SK	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.33%	49.00%	No pass	
31	SK	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-21.58%	49.55%	No pass	
31	SK	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.32%	1.48%	Pass	
31	SK	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.97% S	0.89%	Pass	
31	SK	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.23% S	0.75%	Pass	
31	SK	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.89% S	0.44%	Pass	
31	SK	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.61% S	2.95%	Pass	
31	SK	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.17%	1.69%	Pass	

31	SK	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-55.12% S	30.58%	No pass
31	SK	NO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.64% S	0.23%	Pass
31	SK	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-11.19% S	1.70%	Pass
31	SK	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.96%	72.84%	No pass
31	SK	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.52%	7.56%	
31	SK	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.64% S	1.20%	Pass
31	SK	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-16.81% S	5.50%	No pass
31	SK	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.34%	10.08%	Pass
32	LT	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-11.89% S	21.13%	No pass
32	LT	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.11% S	0.68%	Pass
32	LT	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.62% S	0.33%	Pass
32	LT	HNO3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.49% S	4.00%	Pass
32	LT	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.68%	5.35%	Pass
32	LT	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	25.71% S	11.15%	No pass
32	LT	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.97% S	1.53%	Pass
32	LT	NH3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.46% S	4.17%	Pass
32	LT	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.27% S	3.08%	Pass
32	LT	NO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.56% S	1.84%	Pass
32	LT	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.22% S	1.18%	Pass
32	LT	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.93%	5.06%	
32	LT	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.71% S	2.79%	Pass
32	LT	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.64% S	0.74%	Pass
33	LV	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.22%	1.45%	Pass
33	LV	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-21.26% S	12.98%	Pass
33	LV	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.48%	4.92%	Pass
33	LV	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.98% S	0.37%	Pass
33	LV	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.33% S	1.96%	Pass
33	LV	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.33% S	2.26%	Pass
33	LV	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.67%	65.04%	No pass
33	LV	HNO3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-29.17% S	0.50%	Pass
33	LV	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	10.13% S	2.45%	Pass
33	LV	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.57% S	5.93%	Pass
33	LV	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.83%	0.97%	Pass
33	LV	NH3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.89% S	5.86%	Pass
33	LV	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.35% S	2.53%	Pass
33	LV	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.51% S	7.78%	Pass
33	LV	NO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.71% S	1.68%	Pass
33	LV	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.80% S	2.36%	Pass
33	LV	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.66% S	10.36%	Pass
33	LV	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.64% S	15.95%	
33	LV	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.55%	1.74%	Pass

33	LV	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-19.71% S	2.08%	Pass
33	LV	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.52% S	5.84%	Pass
34	TR	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.57%	16.80%	Pass
34	TR	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	72.83%	92.97%	No pass
34	TR	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2809.35% S	91.55%	No pass
34	TR	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-23.81% S	3.34%	Pass
34	TR	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	18.66% S	31.10%	No pass
34	TR	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	623.93% S	394.05%	No pass
34	TR	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-90.05% S	43.80%	No pass
34	TR	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.06% S	2.53%	Pass
34	TR	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	138.41% S	75.18%	No pass
34	TR	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	58.51%	88.77%	No pass
34	TR	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-22.68% S	5.12%	Pass
34	TR	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-24.35% S	7.50%	Pass
34	TR	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	38.19% S	4.14%	Pass
34	TR	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	316.89%	14596.06%	No pass
34	TR	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.37% S	0.46%	Pass
34	TR	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.20% S	1.44%	Pass
34	TR	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-89.07%	141.39%	No pass
34	TR	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-53.18% S	30.19%	
34	TR	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.89% S	4.55%	Pass
34	TR	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-18.00% S	2.23%	Pass
34	TR	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-25.50% S	20.67%	Pass
35	HR	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	17.03% S	5.68%	Pass
35	HR	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.32% S	2.11%	Pass
35	HR	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	18.02% S	6.68%	Pass
35	HR	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.28% S	0.28%	Pass
35	HR	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.82% S	0.94%	Pass
35	HR	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
35	HR	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
35	HR	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.41% S	1.19%	Pass
35	HR	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.92%	1.56%	Pass
35	HR	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.28% S	0.37%	Pass
35	HR	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.05%	0.56%	Pass
35	HR	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.63% S	2.66%	Pass
35	HR	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.31% S	1.14%	Pass
35	HR	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.86% S	1.89%	Pass
35	HR	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.57% S	0.39%	Pass
35	HR	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.51% S	1.65%	Pass
35	HR	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	20.26% S	5.25%	
35	HR	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.04%	0.78%	Pass

35	HR	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.15% S	0.15%	Pass
35	HR	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
36	SI	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.60%	0.73%	Pass
36	SI	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.71%	2.72%	Pass
36	SI	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.90%	3.28%	Pass
36	SI	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.99% S	0.59%	Pass
36	SI	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.85% S	0.58%	Pass
36	SI	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.96% S	0.91%	Pass
36	SI	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.04%	1.33%	Pass
36	SI	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.55%	2.23%	Pass
36	SI	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.05%	1.45%	Pass
36	SI	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.79% S	0.40%	Pass
36	SI	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.81%	2.50%	Pass
36	SI	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.76% S	0.74%	Pass
36	SI	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.85% S	0.92%	Pass
36	SI	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.87% S	3.30%	Pass
36	SI	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.00%	6.55%	
36	SI	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.76% S	1.04%	Pass
36	SI	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.93% S	3.73%	Pass
38	EE	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	40.61% S	10.87%	Pass
38	EE	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.65% S	0.68%	Pass
38	EE	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.89% S	1.13%	Pass
38	EE	HNO3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.92% S	1.64%	Pass
38	EE	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-11.71% S	5.58%	Pass
38	EE	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.02%	21.21%	No pass
38	EE	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.83% S	2.09%	Pass
38	EE	NH3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.00% S	4.29%	Pass
38	EE	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	29.99% S	6.97%	Pass
38	EE	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.92% S	0.52%	Pass
38	EE	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-15.85% S	6.11%	
38	EE	SO2	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.19% S	1.20%	Pass
38	EE	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.60%	1.78%	Pass
39	PL	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	11.41% S	5.92%	Pass
39	PL	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.38% S	2.11%	Pass
39	PL	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-69.59% S	13.11%	Pass
39	PL	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.23% S	0.00%	Pass
39	PL	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.94%	0.69%	Pass
39	PL	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.92% S	3.39%	Pass
39	PL	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-62.61% S	2.41%	Pass
39	PL	HNO3	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.81% S	0.99%	Pass
39	PL	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.67% S	1.56%	Pass

39	PL	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.60%	4.51%	Pass
39	PL	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.23% S	2.66%	Pass
39	PL	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.30%	20.18%	No pass
39	PL	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.50%	17.25%	No pass
39	PL	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.23%	3.15%	Pass
39	PL	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.91% S	1.83%	Pass
39	PL	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-44.59% S	5.48%	Pass
39	PL	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-65.29% S	7.29%	
39	PL	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.82%	1.79%	Pass
39	PL	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.40%	2.23%	Pass
39	PL	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.84%	10.63%	Pass
41	FR	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.44%	3.38%	Pass
41	FR	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.57%	3.03%	Pass
41	FR	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.30% S	4.64%	Pass
41	FR	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.69% S	5.33%	Pass
41	FR	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.74% S	2.52%	Pass
41	FR	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.55% S	4.49%	Pass
41	FR	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.36%	0.68%	Pass
42	MD	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
42	MD	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-91.07% S	44.07%	No pass
42	MD	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.68% S	2.00%	Pass
42	MD	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
42	MD	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
42	MD	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
42	MD	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
42	MD	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.63% S	2.14%	Pass
42	MD	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-91.14%	69.35%	No pass
42	MD	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-16.37% S	7.06%	
42	MD	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-89.13% S	12.04%	No pass
45	AM	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.57% S	6.34%	Pass
45	AM	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.02%	1.39%	Pass
45	AM	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.92% S	0.94%	Pass
45	AM	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.60%	1.34%	Pass
45	AM	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.22%	1.60%	Pass
45	AM	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-11.07% S	5.08%	Pass
45	AM	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-64.68% S	16.14%	No pass
45	AM	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.91% S	1.96%	Pass
45	AM	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	25.40%	28.75%	
45	AM	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.66% S	1.32%	Pass
46	PL	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
46	PL	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			

46	PL	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.64% S	0.87%	Pass
46	PL	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
46	PL	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
46	PL	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
46	PL	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
46	PL	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
46	PL	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	20.26% S	2.82%	
46	PL	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
47	PL	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.21%	2.18%	Pass
47	PL	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.02%	6.34%	Pass
47	PL	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.04%	3.40%	Pass
47	PL	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.43% S	0.59%	Pass
47	PL	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.32%	1.45%	Pass
47	PL	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.55% S	1.81%	Pass
47	PL	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.93%	10.24%	Pass
47	PL	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.71%	6.25%	Pass
47	PL	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.44% S	0.78%	Pass
47	PL	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.10% S	1.77%	Pass
47	PL	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	22.48% S	2.22%	Pass
47	PL	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.97% S	0.95%	Pass
47	PL	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.65%	1.96%	Pass
47	PL	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.15%	4.42%	Pass
47	PL	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	58.49% S	9.69%	
47	PL	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.68% S	1.19%	Pass
47	PL	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.41% S	0.89%	Pass
48	BE	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.48%	1.09%	Pass
48	BE	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
48	BE	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.04%	3.40%	Pass
48	BE	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.88%	4.24%	Pass
48	BE	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
48	BE	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.79%	5.21%	Pass
48	BE	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	59.86%	163.47%	No pass
48	BE	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
48	BE	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
48	BE	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
48	BE	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.65% S	1.42%	Pass
48	BE	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.10%	1.68%	Pass
48	BE	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.27% S	1.83%	Pass
48	BE	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	10.41% S	4.02%	Pass
48	BE	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
48	BE	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	33.01%	39.96%	No pass

48	BE	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	11.11% S	4.32%	Pass
49	CY	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	34.27% S	16.80%	Pass
49	CY	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.16% S	0.30%	Pass
49	CY	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.64% S	9.08%	Pass
49	CY	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.91% S	0.34%	Pass
49	CY	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
49	CY	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.30% S	2.72%	Pass
49	CY	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.36%	2.08%	Pass
49	CY	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.41%	2.01%	Pass
49	CY	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.05%	0.41%	Pass
49	CY	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.50% S	0.60%	Pass
49	CY	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.21% S	1.64%	Pass
49	CY	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-21.73% S	0.74%	Pass
49	CY	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.53% S	0.26%	Pass
49	CY	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.96% S	1.72%	Pass
49	CY	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
49	CY	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.94%	1.93%	Pass
49	CY	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	44.69% S	22.44%	Pass
50	FR	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.81% S	1.81%	Pass
50	FR	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.24% S	0.19%	Pass
50	FR	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
50	FR	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.45%	1.78%	Pass
50	FR	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.49%	1.12%	Pass
50	FR	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.84% S	0.85%	Pass
50	FR	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	15.84% S	4.42%	Pass
50	FR	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.86% S	3.27%	Pass
50	FR	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
50	FR	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.20%	1.93%	Pass
51	FR	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.36% S	0.73%	Pass
51	FR	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.17% S	2.90%	Pass
51	FR	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.62%	0.68%	Pass
51	FR	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.58% S	0.92%	Pass
51	FR	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	12.84% S	8.20%	Pass
51	FR	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.09% S	0.79%	Pass
51	FR	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	23.02% S	19.04%	Pass
54	PL	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.49%	0.85%	Pass
54	PL	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.45% S	4.23%	Pass
54	PL	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.51%	2.14%	Pass
54	PL	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.75% S	0.19%	Pass
54	PL	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.66% S	4.87%	Pass
54	PL	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.34%	2.38%	Pass

54	PL	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.24%	7.99%	Pass
54	PL	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.51%	2.01%	Pass
54	PL	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.35%	1.16%	Pass
54	PL	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.21% S	2.18%	Pass
54	PL	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.73% S	1.08%	Pass
54	PL	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.36% S	2.63%	Pass
54	PL	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.27% S	0.52%	Pass
54	PL	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.54%	0.59%	Pass
54	PL	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	75.80% S	4.45%	
54	PL	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.41% S	1.63%	Pass
54	PL	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.23%	3.17%	Pass
110	DE	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.52%	9.19%	Pass
110	DE	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.48% S	1.21%	Pass
110	DE	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.25% S	5.04%	Pass
110	DE	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.64% S	1.02%	Pass
110	DE	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.17% S	1.09%	Pass
110	DE	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.73%	3.06%	Pass
110	DE	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.41% S	2.00%	Pass
110	DE	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.74%	0.89%	Pass
110	DE	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.81% S	0.71%	Pass
110	DE	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.46% S	0.40%	Pass
110	DE	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.81% S	2.53%	Pass
110	DE	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.86%	7.99%	Pass
110	DE	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.41%	2.07%	Pass
110	DE	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.91% S	1.05%	Pass
110	DE	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.07% S	1.98%	Pass
110	DE	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.40% S	5.48%	
110	DE	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.23% S	7.13%	No pass
110	DE	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.98% S	8.59%	Pass
112	DE	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.23% S	3.38%	Pass
112	DE	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.21% S	0.75%	Pass
112	DE	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.42% S	1.77%	Pass
112	DE	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.04%	0.96%	Pass
112	DE	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.38% S	0.94%	Pass
112	DE	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.71%	1.92%	Pass
112	DE	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.99% S	0.67%	Pass
112	DE	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.71% S	1.78%	Pass
112	DE	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.57% S	0.75%	Pass
112	DE	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.74% S	0.93%	Pass
112	DE	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.46%	1.39%	Pass
112	DE	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.15% S	0.74%	Pass

112	DE	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.69% S	1.83%	Pass
112	DE	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.44%	708.16%	No pass
112	DE	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	28.82%	24.04%	
112	DE	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.64% S	0.59%	Pass
112	DE	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1055.77% S	521.03%	No pass
114	IT	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	45.18% S	7.24%	Pass
114	IT	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.20% S	0.62%	Pass
114	IT	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.42% S	1.67%	Pass
114	IT	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-14.08% S	3.79%	Pass
114	IT	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	15.77%	11.97%	No pass
114	IT	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.08% S	1.05%	Pass
114	IT	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.83% S	0.83%	Pass
114	IT	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.33% S	0.92%	Pass
114	IT	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.15% S	2.77%	
114	IT	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.42% S	0.89%	Pass
115	DE	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.17% S	3.75%	Pass
115	DE	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.38%	2.41%	Pass
115	DE	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	12.89% S	6.68%	Pass
115	DE	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.83% S	0.74%	Pass
115	DE	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.67% S	0.36%	Pass
115	DE	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.44% S	0.68%	Pass
115	DE	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.54% S	1.33%	Pass
115	DE	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.46%	12.94%	Pass
115	DE	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.56%	2.20%	Pass
115	DE	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.53% S	1.61%	Pass
115	DE	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.22% S	0.53%	Pass
115	DE	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.97% S	0.84%	Pass
115	DE	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.40% S	1.44%	Pass
115	DE	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.13% S	3.04%	Pass
115	DE	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	39.65% S	3.00%	
115	DE	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.82% S	0.89%	Pass
115	DE	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.18% S	1.70%	Pass
120	DE	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.47%	2.78%	Pass
120	DE	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.48% S	5.04%	Pass
120	DE	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.03%	8.45%	Pass
120	DE	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.95% S	0.93%	Pass
120	DE	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	15.76% S	2.40%	Pass
120	DE	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.10% S	2.04%	Pass
120	DE	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.00% S	1.17%	Pass
120	DE	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.10%	4.71%	Pass
120	DE	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.11% S	5.44%	Pass

120	DE	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.95% S	1.49%	Pass
120	DE	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	34.88% S	10.56%	No pass
120	DE	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-96.96% S	24.90%	Pass
120	DE	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.66%	4.45%	Pass
120	DE	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.71% S	4.42%	Pass
120	DE	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.78%	27.28%	
120	DE	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.93% S	4.31%	Pass
120	DE	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	12.39% S	1.46%	Pass
121	DE	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.66% S	9.79%	Pass
121	DE	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-35.32% S	1.81%	Pass
121	DE	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.03% S	0.00%	Pass
121	DE	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.43% S	0.40%	Pass
121	DE	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.37% S	1.31%	Pass
121	DE	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-21.02% S	11.43%	Pass
121	DE	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.57% S	6.00%	Pass
121	DE	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.37% S	2.23%	Pass
121	DE	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-15.59% S	1.45%	Pass
121	DE	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.73% S	1.93%	Pass
121	DE	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.75%	1.67%	Pass
121	DE	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-34.52% S	9.56%	Pass
121	DE	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.30% S	0.52%	Pass
121	DE	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.21% S	2.71%	Pass
121	DE	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.28%	4.72%	
121	DE	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.57% S	1.04%	Pass
121	DE	Zn	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-18.63% S	9.95%	Pass
125	DE	As	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.97% S	2.66%	Pass
125	DE	Ca in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.48% S	2.11%	Pass
125	DE	Cd	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.34% S	3.66%	Pass
125	DE	Cl in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.74% S	0.31%	Pass
125	DE	Cond in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.70% S	0.58%	Pass
125	DE	Cr	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.42%	2.04%	Pass
125	DE	Cu	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.18% S	1.25%	Pass
125	DE	K in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.00%	0.45%	Pass
125	DE	Mg in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.51%	1.60%	Pass
125	DE	Na in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.78% S	1.05%	Pass
125	DE	NH4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.22%	2.22%	Pass
125	DE	Ni	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.76%	3.99%	Pass
125	DE	NO3 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.48% S	0.52%	Pass
125	DE	Pb	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.45%	1.72%	Pass
125	DE	pH in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	24.53% S	5.47%	
125	DE	SO4 in precip	EMEP42	20250125	https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.94% S	0.59%	Pass

125	DE	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.41% S	0.90%	Pass
129	TN	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
129	TN	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	19607.21% S	51.32%	No pass
129	TN	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2236.99% S	1262.26%	No pass
129	TN	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	451.87% S	250.65%	No pass
129	TN	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2715.32% S	38.88%	No pass
129	TN	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	693.36% S	967.72%	No pass
129	TN	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	568.07% S	5424.56%	No pass
145	EE	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.61%	4.23%	Pass
145	EE	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.00% S	3.32%	Pass
145	EE	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	10.59% S	3.28%	Pass
145	EE	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.90% S	0.68%	Pass
145	EE	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.08% S	0.58%	Pass
145	EE	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.14% S	1.02%	Pass
145	EE	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.11% S	2.75%	Pass
145	EE	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.24% S	1.12%	Pass
145	EE	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.15%	3.77%	Pass
145	EE	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.43% S	0.77%	Pass
145	EE	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.37% S	0.67%	Pass
145	EE	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.19% S	2.73%	Pass
145	EE	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.85% S	2.36%	Pass
145	EE	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.03% S	1.19%	Pass
145	EE	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-24.06% S	9.99%	
145	EE	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.88%	2.82%	Pass
145	EE	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.45% S	0.79%	Pass
146	LU	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.62%	0.60%	Pass
146	LU	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.20%	2.02%	Pass
146	LU	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.74% S	1.58%	Pass
146	LU	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.43% S	1.92%	Pass
146	LU	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.51%	1.79%	Pass
146	LU	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.07% S	1.58%	Pass
146	LU	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.67% S	1.05%	Pass
153	SI	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	20.64% S	16.60%	No pass
153	SI	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.98% S	0.96%	Pass
153	SI	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.83% S	2.91%	Pass
153	SI	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.64%	2.45%	Pass
153	SI	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.28% S	11.93%	No pass
153	SI	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.01% S	1.13%	Pass
153	SI	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.46% S	4.75%	Pass
153	SI	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.39% S	1.05%	Pass
153	SI	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-10.50%	8.81%	

153	SI	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.78% S	1.19%	Pass
155	UK	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.83% S	2.41%	Pass
155	UK	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.62% S	1.27%	Pass
155	UK	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.58%	0.80%	Pass
155	UK	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.04%	1.12%	Pass
155	UK	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.81% S	0.34%	Pass
155	UK	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.08% S	0.85%	Pass
155	UK	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-11.65% S	2.81%	Pass
155	UK	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.27%	0.39%	Pass
155	UK	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.99%	5.68%	
155	UK	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.50% S	1.93%	Pass
158	JP	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.25% S	0.60%	Pass
158	JP	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.86% S	0.25%	Pass
158	JP	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.14% S	0.94%	Pass
158	JP	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.88% S	3.33%	Pass
158	JP	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.34% S	0.89%	Pass
158	JP	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.97%	1.90%	Pass
158	JP	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.66%	0.60%	Pass
158	JP	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.46% S	1.14%	Pass
158	JP	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.75% S	1.11%	Pass
158	JP	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.19% S	0.79%	Pass
158	JP	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.70% S	8.01%	
158	JP	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.58% S	0.67%	Pass
158	JP	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.39% S	0.59%	Pass
166	PL	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
166	PL	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.09%	4.23%	Pass
166	PL	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
166	PL	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.73% S	0.49%	Pass
166	PL	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.30% S	0.58%	Pass
166	PL	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
166	PL	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
166	PL	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
166	PL	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.62% S	0.67%	Pass
166	PL	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-7.07% S	1.17%	Pass
166	PL	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.70% S	0.56%	Pass
166	PL	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
166	PL	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.59% S	0.39%	Pass
166	PL	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
166	PL	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	24.53% S	6.61%	
166	PL	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.87% S	0.59%	Pass
166	PL	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.51%	4.12%	Pass

169	UK	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.88% S	3.75%	Pass
169	UK	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.01%	6.18%	Pass
169	UK	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.23%	2.26%	Pass
169	UK	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.21% S	0.42%	Pass
169	UK	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.55% S	1.37%	Pass
169	UK	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.42%	0.66%	Pass
169	UK	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.96%	7.69%	Pass
171	FR	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.41% S	2.54%	Pass
171	FR	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.56% S	2.14%	Pass
171	FR	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.92%	1.58%	Pass
171	FR	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.28% S	2.17%	Pass
171	FR	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.82% S	1.68%	Pass
171	FR	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.29% S	3.76%	Pass
171	FR	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.74% S	4.24%	Pass
178	RU	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.83%	1.93%	Pass
178	RU	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.75%	4.53%	Pass
178	RU	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.73%	15.64%	Pass
178	RU	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.20% S	4.02%	Pass
178	RU	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.42%	3.85%	Pass
178	RU	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.95% S	6.00%	Pass
178	RU	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.16% S	4.33%	Pass
178	RU	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.79%	6.20%	Pass
178	RU	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.65%	6.02%	Pass
178	RU	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.60%	4.51%	Pass
178	RU	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.26% S	1.61%	Pass
178	RU	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.20% S	10.53%	No pass
178	RU	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-25.34% S	8.20%	Pass
178	RU	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.97% S	11.02%	No pass
178	RU	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.35% S	1.70%	Pass
178	RU	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.63%	5.87%	Pass
178	RU	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	30.39%	23.06%	
178	RU	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.48% S	1.40%	Pass
178	RU	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.09% S	5.05%	No pass
178	RU	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.08%	7.01%	Pass
183	RS	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
183	RS	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.23%	13.11%	Pass
183	RS	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.02%	6.79%	Pass
183	RS	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	37.68% S	11.91%	Pass
183	RS	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	10.14% S	11.56%	Pass
183	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.08%	3.24%	Pass
183	RS	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-13.75%	16.56%	Pass

183	RS	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.44%	3.84%	Pass
184	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.25% S	1.84%	Pass
185	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.00%	3.45%	Pass
187	RS	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
187	RS	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
187	RS	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	11.50% S	3.12%	Pass
187	RS	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
187	RS	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
187	RS	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
187	RS	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.59%	18.08%	No pass
187	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	3.69% S	1.38%	Pass
187	RS	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-17.93% S	2.49%	Pass
187	RS	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
187	RS	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
188	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.53%	2.30%	Pass
189	RS	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
189	RS	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
189	RS	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
189	RS	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
189	RS	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
189	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.10% S	2.07%	Pass
189	RS	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
189	RS	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-68.20% S	0.00%	Pass
193	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.01% S	0.69%	Pass
198	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	1.13% S	1.38%	Pass
200	CH	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.88%	4.23%	Pass
200	CH	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.23% S	0.46%	Pass
200	CH	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.05%	0.62%	Pass
200	CH	HNO3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-28.62% S	6.54%	Pass
200	CH	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.55%	1.34%	Pass
200	CH	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	14.61% S	0.41%	Pass
200	CH	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.04% S	0.56%	Pass
200	CH	NH3	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.15% S	3.03%	Pass
200	CH	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	5.64% S	0.56%	Pass
200	CH	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.00% S	0.23%	Pass
200	CH	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.29% S	0.79%	Pass
200	CH	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
200	CH	SO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-28.41% S	7.94%	Pass
200	CH	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.50% S	0.74%	Pass
201	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	14.04% S	2.30%	Pass
203	RS	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-39.19%	25.63%	No pass

203	RS	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	8.10% S	4.23%	Pass
203	RS	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.85%	8.25%	Pass
203	RS	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.61% S	4.81%	Pass
203	RS	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.47% S	4.07%	Pass
203	RS	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	NO2	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	4.00% S	0.92%	Pass
203	RS	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-5.22% S	0.52%	Pass
203	RS	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-57.63% S	0.00%	Pass
203	RS	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
203	RS	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
207	RS	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	12.73% S	5.92%	Pass
207	RS	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-23.36% S	2.72%	Pass
207	RS	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	15.20% S	13.87%	Pass
207	RS	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.27% S	1.27%	Pass
207	RS	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
207	RS	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-2.62%	1.81%	Pass
207	RS	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-48.16% S	3.16%	Pass
207	RS	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.63% S	2.01%	Pass
207	RS	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-8.86% S	1.45%	Pass
207	RS	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.99%	1.05%	Pass
207	RS	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	2.45%	8.58%	No pass
207	RS	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-12.75% S	4.52%	Pass
207	RS	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-9.30% S	2.88%	Pass
207	RS	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-42.55% S	19.60%	Pass
207	RS	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
207	RS	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.16%	4.16%	Pass
207	RS	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-23.22% S	1.95%	Pass
209	RS	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
209	RS	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.28%	5.66%	Pass
209	RS	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-1.45%	5.09%	Pass
209	RS	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	147.67% S	45.96%	No pass
209	RS	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf			
209	RS	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.32% S	3.63%	Pass
209	RS	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	13.21%	19.14%	No pass

209	RS	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-0.13%	6.15%	No pass
209	RS	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-49.67% S	11.06%	
209	RS	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-6.13% S	7.28%	No pass
211	NL	As	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.76% S	2.85%	Pass
211	NL	Ca in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	0.91% S	1.21%	Pass
211	NL	Cd	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.43% S	4.54%	Pass
211	NL	Cl in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-4.33% S	2.58%	Pass
211	NL	Cond in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.83% S	0.40%	Pass
211	NL	Cr	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.40%	6.18%	Pass
211	NL	Cu	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	9.80% S	8.45%	Pass
211	NL	K in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	12.92% S	3.77%	Pass
211	NL	Mg in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	12.62% S	1.08%	Pass
211	NL	Na in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	7.58% S	1.97%	Pass
211	NL	NH4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.93%	6.42%	Pass
211	NL	Ni	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	10.90% S	3.02%	Pass
211	NL	NO3 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.44%	4.93%	Pass
211	NL	Pb	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	6.37% S	2.51%	Pass
211	NL	pH in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-21.47% S	7.04%	
211	NL	SO4 in precip	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	-3.84% S	1.78%	Pass
211	NL	Zn	EMEP42	20250125 https://emep-ccc.nilu.no/static/qameasure/emep42.pdf	17.25% S	5.83%	Pass